

US EPA ARCHIVE DOCUMENT

METHOD 1010A

TEST METHODS FOR FLASH POINT BY PENSKEY-MARTENS CLOSED CUP TESTER

See American Society for Testing Materials (ASTM) Standard D 93-79 or Standard D 93-80 for these method procedures used in the characterization of flash point of liquids. These standards are two of three required method options for determination of the ignitability characteristic in liquid wastes, see 40 CFR 261.21(a)(1). Currently, these ASTM standards can be obtained from Global Engineering Documents, 15 Iverness Way East, Englewood, CO 80112, 1-800-854-7179, <http://global.ihs.com>. At the web site, do a search for "ASTM D 93" (document number) and at the document search results select "view historical revisions." Select the Revision 79 or 80 version.