

SAIC ® An Employee-Owned Company
Science Applications International Corporation

For filing

12/18/92

Changes made to disk are noted on these pages. I did not change the flow chart.

METHOD 8260A

VOLATILE ORGANIC COMPOUNDS BY GAS CHROMATOGRAPHY/MASS SPECTROMETRY (GC/MS):
CAPILLARY COLUMN TECHNIQUE

1.0 SCOPE AND APPLICATION

1.1 Method 8260A is used to determine volatile organic compounds in a variety of solid waste matrices. This method is applicable to nearly all types of samples, regardless of water content, including ground water, aqueous sludges, caustic liquors, acid liquors, waste solvents, oily wastes, mousses, tars, fibrous wastes, polymeric emulsions, filter cakes, spent carbons, spent catalysts, soils, and sediments. The following compounds can be determined by this method:

Analyte	CAS No. ^b	Appropriate Technique	
		Purge-and-Trap	Direct Injection
Acetone	67-64-1	a	a
Acetonitrile	75-05-8	a	a
Acrolein	107-02-8	a	a
Acrylonitrile	107-13-1	a	a
Allyl alcohol	107-18-6	it	a
Allyl chloride	107-05-1	a	a
Benzene	71-43-2	a	a
Benzyl chloride	100-44-7	a	a
Bromoacetone	598-31-2	pp	a
Bromobenzene	108-86-1	a	a
Bromochloromethane	74-97-5	a	a
Bromodichloromethane	75-27-4	a	a
4-Bromofluorobenzene	460-00-4	a	a
Bromoform	75-25-2	a	a
Bromomethane	74-83-9	a	a
iso-Butanol	78-83-1	it	a
n-Butanol	71-36-3	it	a
2-Butanone	78-93-3	a	a
n-Butylbenzene	104-51-8	a	a
sec-Butylbenzene	135-98-8	a	a
tert-Butylbenzene	98-06-6	a	a
Carbon disulfide	75-15-0	a	a
Carbon tetrachloride	6-23-5	a	a
Chloral hydrate	302-17-0	pp	a
Chloroacetone	107-14-2	a it	-
Chlorobenzene	108-90-7	a	a
Chlorobenzene-d ₂	3114-55-4	a	a
2-Chloro-1,3-butadiene	126-99-8	a	a
1-Chlorobutane	109-69-3	a	-
Chlorodibromomethane	124-48-1	a	a
Chloroethane	75-00-3	a	a

Analyte	Appropriate Technique		Direct Injection
	CAS No. ^b	Purge-and-Trap	
2-Chloroethanol	107-03-3	pp	a
bis-(2-chloroethyl) sulfide	505-60-2	pp	a
2-Chloroethyl vinyl ether	110-75-8	a	a
Chloroform	67-66-3	a	a
1-Chlorohexane	544-10-5	a	a
Chloromethane	74-87-3	a	a
Chloroprene	126-99-8	a	pc
3-Chloropropene	75-29-6	a	a
3-Chloropropionitrile	542-76-7	ND	pc
2-Chlorotoluene	95-49-8	a	a
4-Chlorotoluene	106-43-4	a	a
Crotonaldehyde	123-73-9	<i>if it</i>	a
Dibromochloromethane	124-48-1	a	a
1,2-Dibromo-3-chloropropane	96-12-8	pp	a
1,2-Dibromoethane	106-93-4	a	a
Dibromofluoromethane	1868-53-7	a	a
Dibromomethane	74-95-3	a	a
1,2-Dichlorobenzene	95-50-1	a	a
1,2-Dichlorobenzene-d₄		a	a
1,3-Dichlorobenzene	541-73-1	a	a
1,4-Dichlorobenzene	106-46-7	a	a
1,4-Dichlorobenzene-d₄		a	a
cis-1,4-Dichloro-2-butene	1476-11-5	a	a
trans-1,4-Dichloro-2-butene	110-57-6	<i>a PP</i>	a
Dichlorodifluoromethane	75-71-8	a	a
Dichloroethane-d₄		a	a
1,1-Dichloroethane	75-34-3	a	a
1,2-Dichloroethane	107-06-2	a	a
1,1-Dichloroethene	75-35-4	a	a
cis-1,2-Dichloroethene	156-59-2	a	a
trans-1,2-Dichloroethene	156-60-5	a	a
1,2-Dichloropropane	78-87-5	a	a
1,3-Dichloropropane	142-28-9	a	a
2,2-Dichloropropane	594-20-7	a	a
1,3-Dichloro-2-propanol	96-23-1	pp	a
1,1-Dichloropropene	563-58-6	a	a
cis-1,3-Dichloropropene	10061-01-5	a	a
trans-1,3-Dichloropropene	10061-02-6	a	a
1,2,3,4-Diepoxybutane	1464-53-5	a	a
Diethyl ether	60-29-7	a	a
1,4-Difluorobenzene	540-36-3	a	a
1,4-Dioxane	123-91-1	<i>a</i>	a
Epichlorohydrin	106-89-8	i	a
Ethanol	64-17-5	i	a
Ethyl acetate	60-29-7	i	a
Ethylbenzene	100-41-4	a	a
Ethylene oxide	75-21-8	pp	a

Analyte	Appropriate Technique		Direct Injection
	CAS No. ^b	Purge-and-Trap	
Ethyl methacrylate	97-63-2	a	a
Fluorobenzene	462-06-6	a	a
Hexachlorobutadiene	87-68-3	a	a
Hexachloroethane	67-72-1	ND	a
2-Hexanone	591-78-6	a	a
2-Hydroxypropionitrile	78-97-7	ND	pc
Iodomethane	74-88-4	a	a
Isobutyl alcohol	78-83-1	pp	a
Isopropylbenzene	98-82-8	a	a
p-Isopropyltoluene	99-87-6	a	a
Malonitrile	109-77-3	pp	a
Methacrylonitrile	126-98-7	a	a
Methanol	67-56-1	i	ND
Methyl acrylate	96-33-3	a	-
Methyl-t-butyl ether	1634-04-4	a	-
Methylene chloride	75-09-2	a	a
Methyl ethyl ketone	78-93-3	a	a
Methyl iodide	74-88-4	a	a
Methyl methacrylate	80-62-6	a	a
4-Methyl-2-pentanone (MIBK)	108-10-1	a	a
Naphthalene	91-20-3	a	a
Nitrobenzene	98-95-3	a	-
2-Nitropropane	79-46-9	a	-
Pentachloroethane	76-01-7	xi	a pc
Pentafluorobenzene	363-72-4	a	a
2-Picoline	109-06-8	pp	a
Propargyl alcohol	107-19-7	pp	a
6-Propiolactone	57-57-8	pp	a
Propionitrile (ethyl cyanide)	107-12-0	xi	a
n-Propylamine	107-10-8	a	a
n-Propylbenzene	103-65-1	a	a
Pyridine	110-86-1	i	a
Styrene	100-42-5	a	a
1,1,1,2-Tetrachloroethane	630-20-6	a	a
1,1,2,2-Tetrachloroethane	79-34-5	a	a
Tetrachloroethene	127-18-4	a	a
Toluene	108-88-3	a	a
Toluene-d₈		a	a
1,2,3-Trichlorobenzene	87-61-6	a	a
1,2,4-Trichlorobenzene	120-82-1	a	a
1,1,1-Trichloroethane	71-55-6	a	a
1,1,2-Trichloroethane	79-00-5	a	a
Trichloroethene	79-01-6	a	a
Trichlorofluoromethane	75-69-4	a	a
1,2,3-Trichloropropane	96-18-4	a	a
1,2,4-Trimethylbenzene	95-63-6	a	a
1,3,5-Trimethylbenzene	108-67-8	a	a

Analyte	Appropriate Technique		Direct Injection
	CAS No. ^b	Purge-and-Trap	
Vinyl acetate	108-05-4	a	a
Vinyl chloride	75-01-4	a	a
o-Xylene	95-47-6	a	a
m-Xylene	108-38-3	a	a
p-Xylene	106-42-3	a	a

a Adequate response by this technique.

b Chemical Abstract Services Registry Number.

pp Poor purging efficiency resulting in high EQLs.

i Inappropriate technique for this analyte.

pc Poor chromatographic behavior.

ND Not detected

method analysis only when purged at 20°C

1.2 Method 8260A can be used to quantitate most volatile organic compounds that have boiling points below 200°C and that are insoluble or slightly soluble in water. Volatile water-soluble compounds can be included in this analytical technique. However, for the more soluble compounds, quantitation limits are approximately ten times higher because of poor purging efficiency. Such compounds include low-molecular-weight halogenated hydrocarbons, aromatics, ketones, nitriles, acetates, acrylates, ethers, and sulfides. See Tables 1 and 2 for lists of analytes and retention times that have been evaluated on a purge-and-trap GC/MS system. Also, the method detection limits for 25 mL sample volumes are presented.

1.3 The estimated quantitation limit (EQL) of Method 8260A for an individual compound is somewhat instrument dependent. Using standard quadrupole instrumentation, limits should be approximately 5 µg/kg (wet weight) for soil/sediment samples, 0.5 mg/kg (wet weight) for wastes, and 5 µg/L for ground water (see Table 3). Somewhat lower limits may be achieved using an ion trap mass spectrometer or other instrumentation of improved design. No matter which instrument is used, EQLs will be proportionately higher for sample extracts and samples that require dilution or reduced sample size to avoid saturation of the detector.

1.4 Method 8260A is based upon a purge-and-trap, gas chromatographic/mass spectrometric (GC/MS) procedure. This method is restricted to use by, or under the supervision of, analysts experienced in the use of purge-and-trap systems and gas chromatograph/mass spectrometers, and skilled in the interpretation of mass spectra and their use as a quantitative tool.

1.5 An additional method for sample introduction is direct injection. This technique has been tested for the analysis of waste oil diluted with hexadecane 1:1 (vol/vol) and may have application for the analysis of some alcohols and aldehydes in aqueous samples.

7.1.1 Direct injection - in very limited application, (e.g., volatiles in waste oil or aqueous process wastes) direct injection of aqueous samples or samples diluted according to Method 3585 may be appropriate. Direct injection has been used for the analysis of volatiles in waste oil (diluted 1:1 with hexadecane) and for determining if the sample is ignitable (aqueous injection, Methods 1010 or 1020). Direct injection is only permitted for the determination of volatiles at the TCLP regulatory limits, at concentrations in excess of 10,000 $\mu\text{g/L}$, or for water-soluble compounds that do not purge.

7.1.2 Purge-and-trap for aqueous samples, see Method 5030 for details.

7.1.3 Purge-and-trap for solid samples, see Method ⁵⁰³⁰5035 for details.

7.2 Chromatographic conditions

7.2.1 General:

Injector temperature: 200-225°C
Transfer line temperature: 250-300°C

7.2.2 Column 1 (A sample chromatogram is presented in Figure 5)

Carrier gas (He) flow rate: 15 mL/min
Initial temperature: 10°C, hold for 5 minutes
Temperature program: 6°C/min to 160°C
Final temperature: 160°C, hold until all expected compounds have eluted.

7.2.3 Column 2, Cryogenic cooling (A sample chromatogram is presented in Figure 6)

Carrier gas (He) flow rate: 15 mL/min
Initial temperature: 10°C, hold for 5 minutes
Temperature program: 6°C/min to 160°C
Final temperature: 160°C, hold until all expected compounds have eluted.

7.2.4 Column 2, Non-cryogenic cooling (A sample chromatogram is presented in Figure 7). It is recommended that carrier gas flow and split and make-up gases be set using performance of standards as guidance. Set the carrier gas head pressure to \approx 10 psi and the split to \approx 30 mL/min. Optimize the make-up gas flow for the separator (approximately 30 mL/min) by injecting BFB, and determining the optimum response when varying the make-up gas. This will require several injections of BFB. Next, make several injections of the volatile working standard with all analytes of interest. Adjust the carrier and split to provide optimum chromatography and response. This is an especially critical adjustment for the volatile gas analytes. The head pressure should optimize between 8-12 psi and the split between 20-60 mL/min. The use of the splitter is important to minimize the effect of water on analyte response, to allow the use of a larger volume of helium during trap desorption, and to slow column flow.

Initial temperature: 45°C, hold for 2 minutes
Temperature program: 8°C/min to 200°C
Final temperature: 200°C, hold for 6 minutes.

A trap preheated to 150°C prior to trap desorption is required to provide adequate chromatography of the gas analytes.

7.2.5 Column 3 (A sample chromatogram is presented in Figure 8)

Carrier gas (He) flow rate: 4 mL/min
Initial temperature: 10°C, hold for 5 minutes
Temperature program: 6°C/min to 70°C, then 15°C/min to 145°C
Final temperature: 145°C, hold until all expected compounds have eluted.

7.2.6 Direct injection - Column 4

Carrier gas (He) flow rate: 4 mL/min
Column: J&W DB-24, 70m x 0.53 mm
Initial temperature: 40°C, hold for 3 minutes
Temperature program: 8°C/min
Final temperature: 260°C, hold until all expected compounds have eluted.
Column Bake out (direct inj): 75 minutes
Injector temperature: 200-225°C
Transfer line temperature: 250-300°C

7.3 Initial calibration - the recommended MS operating conditions

Mass range: 35-260 amu
Scan time: 0.6-2 sec/scan
Source temperature: According to manufacturer's specifications
Ion trap only: Axial modulation 4.0 volts
Manifold set 220°C
Emission current 30 amps

~~Set up purge & trap~~
7.3.1 Each GC/MS system must be hardware-tuned to meet the criteria in Table 4 for a 5-50 ng injection or purging of 4-bromofluorobenzene (2 µL injection of the BFB standard). Analyses must not begin until these criteria are met.

~~Set up purge & trap~~
7.3.2 A set of at least five calibration standards containing the method analytes is needed. One calibration standard should contain each analyte at a concentration approaching but greater than the method detection limit (Table 1) for that compound; the other calibration standards should contain analytes at concentrations that define the range of the method. Calibration should be done using the sample introduction technique that will be used for samples. For Method 5030, the purging efficiency for 5 mL of water is greater than for 25 mL. Therefore, develop the standard curve with whichever volume of sample that will be analyzed.

7.3.2.1 To prepare a calibration standard for purge-and-trap or aqueous direct injection, add an appropriate volume of a secondary dilution standard solution to an aliquot of organic-free reagent water in a volumetric flask. Use a microsyringe and rapidly inject the alcoholic standard into the expanded area of the filled volumetric flask. Remove the needle as quickly as possible after injection. Mix by inverting the flask three times only. Discard the contents contained in the neck of the flask. Aqueous standards are not stable and should be prepared daily. Transfer 5.0 mL (or 25 mL if lower detection limits are required) of each standard to a gas tight syringe along with 10 μ L of internal standard. Then transfer the contents to a purging device or syringe. *Perform purge*

7.3.2.2 To prepare a calibration standard for direct injection analysis of oil, dilute standards in hexadecane.

7.3.3 Tabulate the area response of the characteristic ions (see Table 5) against concentration for each compound and each internal standard. Calculate response factors (RF) for each compound relative to one of the internal standards. The internal standard selected for the calculation of the RF for a compound should be the internal standard that has a retention time closest to the compound being measured (Section 7.6.2). The RF is calculated as follows:

$$RF = (A_x C_{is}) / (A_{is} C_x)$$

where:

A_x = Area of the characteristic ion for the compound being measured.
 A_{is} = Area of the characteristic ion for the specific internal standard.
 C_{is} = Concentration of the specific internal standard.
 C_x = Concentration of the compound being measured.

7.3.4 The average RRF must be calculated and recorded for each compound. A system performance check should be made before this calibration curve is used. Five compounds (the System Performance Check Compounds, or SPCCs) are checked for a minimum average relative response factor. These compounds are chloromethane; 1,1-dichloroethane; bromoform; 1,1,2,2-tetrachloroethane; and chlorobenzene. ~~The minimum acceptable average RRF for these compounds should be 0.100.~~ These compounds are used to check compound instability and to check for degradation caused by contaminated lines or active sites in the system. Examples of these occurrences are:

7.3.4.1 Chloromethane - This compound is the most likely compound to be lost if the purge flow is too fast.

7.3.4.2 Bromoform - This compound is one of the compounds most likely to be purged very poorly if the purge flow is too slow. Cold spots and/or active sites in the transfer lines may adversely affect response. Response of the quantitation ion (m/z 173) is

alternative to average response factor calibration, and a useful diagnostic of standard preparation accuracy and absorption activity in the chromatographic system.

7.3.7 These curves are verified each shift by purging a performance standard. Recalibration is required only if calibration and on-going performance criteria cannot be met.

7.4 ~~Daily~~ GC/MS calibration *verification following method 5030*

7.4.1 Prior to the analysis of samples, inject or purge 5-50 ng of the 4-bromofluorobenzene standard. The resultant mass spectra for the BFB must meet all of the criteria given in Table 4 before sample analysis begins. These criteria must be demonstrated each 12-hour shift.

7.4.2 The initial calibration curve (Section 7.3) for each compound of interest must be checked and verified once every 12 hours during analysis with the introduction technique used for samples. This is accomplished by analyzing a calibration standard that is at a concentration near the midpoint concentration for the working range of the GC/MS by checking the SPCC and CCC.

7.4.3 System Performance Check Compounds (SPCCs) - A system performance check must be made each 12 hours. If the SPCC criteria are met, a comparison of relative response factors is made for all compounds. This is the same check that is applied during the initial calibration. If the minimum relative response factors are not met, the system must be evaluated, and corrective action must be taken before sample analysis begins. Some possible problems are standard mixture degradation, injection port inlet contamination, contamination at the front end of the analytical column, and active sites in the column or chromatographic system.

7.4.3.1 The minimum relative response factor for volatile SPCCs are as follows:

Chloromethane	0.1
1,1-Dichloroethane	0.1
Bromoform	0.25
Chlorobenzene	0.3
1,1,2,2-Tetrachloroethane	0.3

7.4.4 Calibration Check Compounds (CCCs) - After the system performance check is met, CCCs listed in Section 7.3.8 are used to check the validity of the initial calibration.

Calculate the percent drift using the following equation:

$$\% \text{ Drift} = (C_i - C_c) / C_i \times 100$$

where:

C_I = Calibration Check Compound standard concentration.
 C_c = Measured concentration using selected quantitation method.

If the percent drift for each CCC is less than 20%, the initial calibration is assumed to be valid. If the criterion is not met (> 20% drift), for any one CCC, corrective action must be taken. Problems similar to those listed under SPCCs could affect this criterion. If no source of the problem can be determined after corrective action has been taken, a new five point calibration MUST be generated. This criterion MUST be met before quantitative sample analysis begins. If the CCCs are not required analytes by the permit, then all required analytes must meet the 20% drift criterion.

7.4.5 The internal standard responses and retention times in the check calibration standard must be evaluated immediately after or during data acquisition. If the retention time for any internal standard changes by more than 30 seconds from the last check calibration (12 hours), the chromatographic system must be inspected for malfunctions and corrections must be made, as required. If the EICP area for any of the internal standards changes by a factor of two (-50% to +100%) from the last daily calibration standard check, the mass spectrometer must be inspected for malfunctions and corrections must be made, as appropriate. When corrections are made, reanalysis of samples analyzed while the system was malfunctioning are necessary.

7.4.6 For compounds which exhibit linearity of response, the RRF of the daily check standard may be used for quantitation, provided the criteria for SPCCs and CCCs are satisfied.

7.5 GC/MS analysis

7.5.1 It is highly recommended that the extract be screened on a headspace-GC/FID (Methods 3810/8015), headspace-GC/PID/ELCD (Methods 3810/8021), or waste dilution-GC/PID/ELCD (Methods 3585/8021) using the same type of capillary column. This will minimize contamination of the GC/MS system from unexpectedly high concentrations of organic compounds. Use of screening is particularly important when this method is used to achieve low detection levels.

7.5.2 All samples and standard solutions must be allowed to warm to ambient temperature before analysis. *Set up of purge & trap*

7.5.3 BFB tuning criteria and daily GC/MS calibration *verification* criteria must be met before analyzing samples.

7.5.3.1 Remove the plunger from a 5 mL syringe and attach a closed syringe valve. If lower detection limits are required, use a 25 mL syringe. Open the sample or standard bottle, which has been allowed to come to ambient temperature, and carefully pour the sample

7.5.4.2.5 If less than five samples are used for compositing, a proportionately smaller syringe may be used unless a 25 mL sample is to be purged.

7.5.5 Add 10.0 μL of surrogate spiking solution and 10 μL of internal standard spiking solution to each sample. The surrogate and internal standards may be mixed and added as a single spiking solution. The addition of 10 μL of the surrogate spiking solution to 5 mL of sample is equivalent to a concentration of 50 $\mu\text{g/L}$ of each surrogate standard. The addition of 10 μL of the surrogate spiking solution to 5 g of sample is equivalent to a concentration of 50 $\mu\text{g/kg}$ of each surrogate standard.

7.5.5.1 If a more sensitive mass spectrometer is employed to achieve lower detection levels, more dilute surrogate and internal standard solutions may be required.

Perform purge and trap
7.5.6 If the initial analysis of sample or a dilution of the sample has a concentration of analytes that exceeds the initial calibration range, the sample must be reanalyzed at a higher dilution. Secondary ion quantitation is allowed only when there are sample interferences with the primary ion. When a sample is analyzed that has saturated ions from a compound, this analysis must be followed by a blank organic-free reagent water analysis. If the blank analysis is not free of interferences, the system must be decontaminated. Sample analysis may not resume until the blank analysis is demonstrated to be free of interferences.

7.5.6.1. All dilutions should keep the response of the major constituents (previously saturated peaks) in the upper half of the linear range of the curve. Proceed to Sections 7.6.1 and 7.6.2 for qualitative and quantitative analysis.

7.5.7 For matrix spike analysis, add 10 μL of the matrix spike solution (Section 5.13) to the 5 mL of sample to be purged. Disregarding any dilutions, this is equivalent to a concentration of 50 $\mu\text{g/L}$ of each matrix spike standard.

7.6 Data interpretation

7.6.1 Qualitative analysis

7.6.1.1 The qualitative identification of compounds determined by this method is based on retention time, and on comparison of the sample mass spectrum, after background correction, with characteristic ions in a reference mass spectrum. The reference mass spectrum must be generated by the laboratory using the conditions of this method. The characteristic ions from the reference mass spectrum are defined to be the three ions of greatest relative intensity, or any ions over 30% relative intensity if less than three such ions occur in the reference spectrum. Compounds should be identified as present when the criteria below are met.

7.6.1.1.1 The intensities of the characteristic ions of a compound maximize in the same scan or within one scan of

TABLE 10
DIRECT INJECTION ANALYSIS OF NEW OIL AT 5 PPM

<u>Compound</u>	<u>Recovery (%)</u>	<u>%RSD</u>	<u>Blank (ppm)</u>	<u>Spike</u>
Acetone	91	14.8	1.9	5.0
Benzene	86	21.3	0.1	0.5
n-Butanol*,**	107	27.8	0.5	5.0
iso-Butanol*,**	95	19.5	0.9	5.0
Carbon tetrachloride	86	44.7	0.0	0.5
Carbon disulfide**	53	22.3	0.0	5.0
Chlorobenzene	81	29.3	0.0	5.0
Chloroform	84	29.3	0.0	6.0
1,4-Dichlorobenzene	98	24.9	0.0	7.5
1,2-Dichloroethane	101	23.1	0.0	0.5
1,1-Dichloroethene	97	45.3	0.0	0.7
Diethyl ether	76	24.3	0.0	5.0
Ethyl acetate	113	27.4	0.0	5.0
Ethylbenzene	83	30.1	0.2	5.0
Hexachloroethane	71	30.3	0.0	3.0
Methylene chloride	98	45.3	0.0	5.0
Methyl ethyl ketone	79	24.6	0.4	5.0
MIBK	93	31.4	0.0	5.0
Nitrobenzene	89	30.3	0.0	2.0
Pyridine	31	35.9	0.0	5.0
Tetrachloroethene	82	27.1	0.0	0.7
Trichlorofluoromethane	76	27.6	0.0	5.0
1,1,2-Cl ₃ F ₃ ethane	69	29.2	0.0	5.0
Toluene	73	21.9	0.6	5.0
Trichloroethene	66	28.0	0.0	0.5
Vinyl chloride	63	35.2	0.0	0.2
o-Xylene	83	29.5	0.4	5.0
m/p-Xylene	84	29.5	0.6	10.0

* Alternate mass employed

** IS quantitation

METHOD 8260A

VOLATILE ORGANIC COMPOUNDS BY GAS CHROMATOGRAPHY/MASS
SPECTROMETRY (GC/MS):
CAPILLARY COLUMN TECHNIQUE

1.0 SCOPE AND APPLICATION

1.1 Method 8260A is used to determine volatile organic compounds in a variety of solid waste matrices. This method is applicable to nearly all types of samples, regardless of water content, including ground water, aqueous sludges, caustic liquors, acid liquors, waste solvents, oily wastes, mousses, tars, fibrous wastes, polymeric emulsions, filter cakes, spent carbons, spent catalysts, soils, and sediments. The following compounds can be determined by this method:

Analyte	CAS No. ^b	Appropriate Technique	
		Purge-and-Trap	Direct Injection
Acetone	67-64-1	ppa	a
Acetonitrile	75-05-8	ppa	a
Acrolein	107-02-8	ppa	a
Acrylonitrile	107-13-1	ppa	a
Allyl alcohol	107-18-6	hti	a
Allyl chloride	107-05-1	a	a
Benzene	71-43-2	a	a
Benzyl chloride	100-44-7	a	a
Bromoacetone	598-31-2	pp	a
Bromobenzene	108-86-1	a	a
Bromochloromethane	74-97-5	a	a
Bromodichloromethane	75-27-4	a	a
4-Bromofluorobenzene	460-00-4	a	a
Bromoform	75-25-2	a	a
Bromomethane	74-83-9	a	a
iso-Butanol	78-83-1	hti	a
n-Butanol	71-36-3	hti	a
2-Butanone	78-93-3	ppa	a
n-Butylbenzene	104-51-8	a	a
sec-Butylbenzene	135-98-8	a	a

tert-Butylbenzene	98-06-6	a	a
Carbon disulfide	75-15-0	ppa	a
Carbon tetrachloride	6-23-5	a	a
Chloral hydrate	302-17-0	pp	a
Chloroacetonitrile	107-14-2	hta	-
Chlorobenzene	108-90-7	a	a
Chlorobenzene-d₅	3114-55-4	a	a
2-Chloro-1,3-butadiene	126-99-8	a	a
1-Chlorobutane	109-69-3	a	-
Chlorodibromomethane	124-48-1	a	a
Chloroethane	75-00-3	a	a

Appropriate Technique

Analyte	CAS No. ^b	Direct	
		Purge-and-Trap	Injection
2-Chloroethanol	107-03-3	pp	a
bis-(2-chloroethyl) sulfide	505-60-2	pp	a
2-Chloroethyl vinyl ether	110-75-8	a	a
Chloroform	67-66-3	a	a
1-Chlorohexane	544-10-5	a	a
Chloromethane	74-87-3	a	a
Chloroprene	126-99-8	a	pc
3-Chloropropene	75-29-6	a	a
3-Chloropropionitrile	542-76-7	ND	pc
2-Chlorotoluene	95-49-8	a	a
4-Chlorotoluene	106-43-4	a	a
Crotonaldehyde	123-73-9	hti	a
Dibromochloromethane	124-48-1	a	a
1,2-Dibromo-3-chloropropane	96-12-8	pp	a
1,2-Dibromoethane	106-93-4	a	a
Dibromofluoromethane	1868-53-7	a	a
Dibromomethane	74-95-3	a	a
1,2-Dichlorobenzene	95-50-1	a	a
1,2-Dichlorobenzene-d₄		a	a
1,3-Dichlorobenzene	541-73-1	a	a
1,4-Dichlorobenzene	106-46-7	a	a
1,4-Dichlorobenzene-d₄		a	a
cis-1,4-Dichloro-2-butene	1476-11-5	a	a
trans-1,4-Dichloro-2-butene	110-57-6	ppa	a
Dichlorodifluoromethane	75-71-8	a	a
Dichloroethane-d₄		a	a
1,1-Dichloroethane	75-34-3	a	a
1,2-Dichloroethane	107-06-2	a	a
1,1-Dichloroethene	75-35-4	a	a

cis-1,2-Dichloroethene	156-59-2		a		a
trans-1,2-Dichloroethene	156-60-5	a			a
1,2-Dichloropropane	78-87-5	a			a
1,3-Dichloropropane	142-28-9		a		a
2,2-Dichloropropane	594-20-7		a		a
1,3-Dichloro-2-propanol	96-23-1		pp		a
1,1-Dichloropropene	563-58-6		a		a
cis-1,3-Dichloropropene	10061-01-5		a		a
trans-1,3-Dichloropropene	10061-02-6		a		a
1,2,3,4-Diepoxybutane	1464-53-5		a		a
Diethyl ether	60-29-7		a		a
1,4-Difluorobenzene	540-36-3		a		a
1,4-Dioxane	123-91-1		ppa		a
Epichlorohydrin	106-89-8		i		a
Ethanol	64-17-5		i		a
Ethyl acetate	60-29-7		i		a
Ethylbenzene	100-41-4		a		a
Ethylene oxide	75-21-8		pp		a

Appropriate Technique

Analyte	CAS No. ^b	Direct	
		Purge-and-Trap	Injection
Ethyl methacrylate	97-63-2	a	a
Fluorobenzene	462-06-6	a	a
Hexachlorobutadiene	87-68-3	a	a
Hexachloroethane	67-72-1	ND	a
2-Hexanone	591-78-6	ppa	a
2-Hydroxypropionitrile	78-97-7	ND	pc
Iodomethane	74-88-4	a	a
Isobutyl alcohol	78-83-1	pp	a
Isopropylbenzene	98-82-8	a	a
p-Isopropyltoluene	99-87-6	a	a
Malononitrile	109-77-3	pp	a
Methacrylonitrile	126-98-7	ppa	a
Methanol	67-56-1	i	ND
Methyl acrylate	96-33-3	a	-
Methyl-t-butyl ether	1634-04-4	a	-
Methylene chloride	75-09-2	a	a
Methyl ethyl ketone	78-93-3	a	a
Methyl iodide	74-88-4	a	a
Methyl methacrylate	80-62-6	a	a
4-Methyl-2-pentanone (MIBK)	108-10-1	ppa	a
Naphthalene	91-20-3	a	a
Nitrobenzene	98-95-3	a	-

2-Nitropropane	79-46-9	a	-
Pentachloroethane	76-01-7	ia	a
Pentafluorobenzene	363-72-4	a	a
2-Picoline	109-06-8	pp	a
Propargyl alcohol	107-19-7	pp	a
-Propiolactone	57-57-8	pp	a
Propionitrile (ethyl cyanide)		107-12-0	hta
pca			
n-Propylamine	107-10-8	a	a
n-Propylbenzene	103-65-1	a	a
Pyridine	110-86-1	i	a
Styrene	100-42-5	a	a
1,1,1,2-Tetrachloroethane	630-20-6	a	a
1,1,2,2-Tetrachloroethane	79-34-5	a	a
Tetrachloroethene	127-18-4	a	a
Toluene	108-88-3	a	a
Toluene-d₈		a	a
1,2,3-Trichlorobenzene	87-61-6	a	a
1,2,4-Trichlorobenzene	120-82-1	a	a
1,1,1-Trichloroethane	71-55-6	a	a
1,1,2-Trichloroethane	79-00-5	a	a
Trichloroethene	79-01-6	a	a
Trichlorofluoromethane	75-69-4	a	a
1,2,3-Trichloropropane	96-18-4	a	a
1,2,4-Trimethylbenzene	95-63-6	a	a
1,3,5-Trimethylbenzene	108-67-8	a	a

Appropriate Technique

Analyte	CAS No. ^b	Direct	
		Purge-and-Trap	Injection
Vinyl acetate	108-05-4	a	a
Vinyl chloride	75-01-4	a	a
o-Xylene	95-47-6	a	a
m-Xylene	108-38-3	a	a
p-Xylene	106-42-3	a	a

- a Adequate response by this technique.
b Chemical Abstract Services Registry Number.
pp Poor purging efficiency resulting in high EQLs.
i Inappropriate technique for this analyte.
pc Poor chromatographic behavior.
ND Not detected.

ht Method analyte only when purged at 80°C.

1.2 Method 8260A can be used to quantitate most volatile organic compounds that have boiling points below 200°C and that are insoluble or slightly soluble in water. Volatile water-soluble compounds can be included in this analytical technique. However, for the more soluble compounds, quantitation limits are approximately ten times higher because of poor purging efficiency. Such compounds include low-molecular-weight halogenated hydrocarbons, aromatics, ketones, nitriles, acetates, acrylates, ethers, and sulfides. See Tables 1 and 2 for lists of analytes and retention times that have been evaluated on a purge-and-trap GC/MS system. Also, the method detection limits for 25 mL sample volumes are presented.

1.3 The estimated quantitation limit (EQL) of Method 8260A for an individual compound is somewhat instrument dependent. Using standard quadrupole instrumentation, limits should be approximately 5 μ g/kg (wet weight) for soil/sediment samples, 0.5 mg/kg (wet weight) for wastes, and 5 μ g/L for ground water (see Table 3). Somewhat lower limits may be achieved using an ion trap mass spectrometer or other instrumentation of improved design. No matter which instrument is used, EQLs will be proportionately higher for sample extracts and samples that require dilution or reduced sample size to avoid saturation of the detector.

1.4 Method 8260A is based upon a purge-and-trap, gas chromatographic/mass spectrometric (GC/MS) procedure. This method is restricted to use by, or under the supervision of, analysts experienced in the use of purge-and-trap systems and gas chromatograph/mass spectrometers, and skilled in the interpretation of mass spectra and their use as a quantitative tool.

1.5 An additional method for sample introduction is direct injection. This technique has been tested for the analysis of waste oil diluted with hexadecane 1:1 (vol/vol) and may have application for the analysis of some alcohols and aldehydes in aqueous samples.

2.0 SUMMARY OF METHOD

2.1 The volatile compounds are introduced into the gas chromatograph by the purge-and-trap method or by direct injection (in limited applications). Purged sample components are trapped in a tube containing suitable sorbent materials. When purging is complete, the sorbent tube is heated and backflushed with helium to desorb trapped sample components. The analytes are desorbed directly to a large bore capillary or cryofocussed on a capillary precolumn before being flash evaporated to a narrow bore capillary

for analysis. The column is temperature programmed to separate the analytes which are then detected with a mass spectrometer (MS) interfaced to the gas chromatograph. Wide bore capillary columns require a jet separator, whereas narrow bore capillary columns can be directly interfaced to the ion source.

2.2 If the above sample introduction techniques are not applicable, a portion of the sample is dispersed in solvent to dissolve the volatile organic constituents. A portion of the solution is combined with organic-free reagent water in the purge chamber. It is then analyzed by purge-and-trap GC/MS following the normal water method.

2.3 Analytes eluted from the capillary column are introduced into the mass spectrometer via a jet separator or a direct connection. Identification of target analytes is accomplished by comparing their mass spectra with the electron impact (or electron impact-like) spectra of authentic standards. Quantitation is accomplished by comparing the response of a major (quantitation) ion relative to an internal standard with a five-point calibration curve.

2.4 The method includes specific calibration and quality control steps that replace the general requirements in Method 8000.

3.0 INTERFERENCES

3.1 Major contaminant sources are volatile materials in the laboratory and impurities in the inert purging gas and in the sorbent trap. The use of non-polytetrafluoroethylene (PTFE) thread sealants, plastic tubing, or flow controllers with rubber components should be avoided since such materials out-gas organic compounds which will be concentrated in the trap during the purge operation. Analyses of calibration and reagent blanks provide information about the presence of contaminants. When potential interfering peaks are noted in blanks, the analyst should change the purge gas source and regenerate the molecular sieve purge gas filter (Figure 1). Subtracting blank values from sample results is not permitted. If reporting values not corrected for blanks result in what the laboratory feels is a false positive for a sample, this should be fully explained in text accompanying the uncorrected data.

3.2 Interfering contamination may occur when a sample containing low concentrations of volatile organic compounds is analyzed immediately after a sample containing high concentrations of volatile organic compounds. The preventive technique is rinsing of the purging apparatus and sample syringes with two portions of organic-free reagent water between samples. After analysis of a

sample containing high concentrations of volatile organic compounds, one or more calibration blanks should be analyzed to check for cross contamination. For samples containing large amounts of water soluble materials, suspended solids, high boiling compounds or high concentrations of compounds being determined, it may be necessary to wash the purging device with a soap solution, rinse it with organic-free reagent water, and then dry the purging device in an oven at 105°C. In extreme situations, the whole purge and trap device may require dismantling and cleaning. Screening of the samples prior to purge and trap GC/MS analysis is highly recommended to prevent contamination of the system. This is especially true for soil and waste samples. Screening may be accomplished with an automated headspace technique or by Method 3820 (Hexadecane Extraction and Screening of Purgeable Organics).

3.2.1 The low purging efficiency of many analytes from a 25 mL sample often results in significant concentrations remaining in the sample purge vessel after analysis. After removal of the analyzed sample aliquot and three rinses of the purge vessel with analyte free water, it is required that the empty vessel be subjected to a heated purge cycle prior to the analysis of another sample in the same purge vessel to reduce sample to sample carryover.

3.3 Special precautions must be taken to analyze for methylene chloride. The analytical and sample storage area should be isolated from all atmospheric sources of methylene chloride. Otherwise random background levels will result. Since methylene chloride will permeate through PTFE tubing, all gas chromatography carrier gas lines and purge gas plumbing should be constructed from stainless steel or copper tubing. Laboratory clothing worn by the analyst should be clean since clothing previously exposed to methylene chloride fumes during liquid/liquid extraction procedures can contribute to sample contamination.

3.4 Samples can be contaminated by diffusion of volatile organics (particularly methylene chloride and fluorocarbons) through the septum seal into the sample during shipment and storage. A trip blank prepared from organic-free reagent water and carried through the sampling and handling protocol can serve as a check on such contamination.

3.5 Use of sensitive mass spectrometers to achieve lower detection level will increase the potential to detect laboratory contaminants as interferences.

3.6 Direct injection - Some contamination may be eliminated by baking out the column between analyses. Changing the injector liner will reduce the potential for cross-contamination. A portion of the analytical column may need to be removed in the case of

extreme contamination. Use of direct injection will result in the need for more frequent instrument maintenance.

3.7 If hexadecane is added to samples or petroleum samples are analyzed, some chromatographic peaks will elute after the target analytes. The oven temperature program must include a post-analysis bake out period to ensure that semi-volatile hydrocarbons are volatilized.

4.0 APPARATUS AND MATERIALS

4.1 Purge-and-trap device - aqueous samples, described in Method 5030.

4.2 Purge-and-trap device - solid samples, described in Method 5030.

4.3 Injection port liners (HP catalogue #18740-80200, or equivalent) are  analysis by placing a 1-cm plug of  pyrex wool approximately 50-60 mm down the length of the injection port towards the oven. An 0.53 mm id column is mounted 1 cm into the liner from the oven side of the injection port, according to manufacturer's specifications.

4.4 Gas chromatography/mass spectrometer/data system

4.4.1 Gas chromatograph - An analytical system complete with a temperature-programmable gas chromatograph suitable for splitless injection or interface to purge-and-trap apparatus. The system includes all required accessories, including syringes, analytical columns, and gases. The GC should be equipped with variable constant differential flow controllers so that the column flow rate will remain constant throughout desorption and temperature program operation. For some column configurations, the column oven must be cooled to $< 30^{\circ}\text{C}$, therefore, a subambient oven controller may be required. The capillary column should be directly coupled to the source.

4.4.1.1 Capillary precolumn interface when using cryogenic cooling - This device interfaces the purge and trap concentrator to the capillary gas chromatograph. The interface condenses the desorbed sample components and focuses them into a narrow band on an uncoated fused silica capillary precolumn. When the interface is flash heated, the sample is transferred to the analytical capillary column.

4.4.1.1.1 During the cryofocussing step, the temperature of the fused silica in the interface is maintained at -150°C under a stream of liquid nitrogen. After the desorption period, the interface must be capable of rapid heating to 250°C in 15 seconds or less to complete the transfer of analytes.

4.4.2 Gas chromatographic columns

4.4.2.1 Column 1 - 60 m x 0.75 mm ID capillary column coated with VOCOL (Supelco), 1.5 μm film thickness, or equivalent.

4.4.2.2 Column 2 - 30 - 75 m x 0.53 mm ID capillary column coated with DB-624 (J&W Scientific), Rt_x -502.2 (RESTEK), or VOCOL (Supelco), 3 μm film thickness, or equivalent.

4.4.2.3 Column 3 - 30 m x 0.25 - 0.32 mm ID capillary column coated with 95% dimethyl - 5% diphenyl polysiloxane (DB-5, Rt_x -5, SPB-5, or equivalent), 1 μm film thickness.

4.4.3 Mass spectrometer - Capable of scanning from 35 to 300 amu every 2 sec or less, using 70 volts (nominal) electron energy in the electron impact ionization mode. The mass spectrometer must be capable of producing a mass spectrum for p-Bromofluorobenzene (BFB) which meets all of the criteria in Table 4 when 5-50 ng of the GC/MS tuning standard (BFB) is injected through the GC. To ensure sufficient precision of mass spectral data, the desirable MS scan rate allows acquisition of at least five spectra while a sample component elutes from the GC.

4.4.3.1 The ion trap mass spectrometer may be used if it is capable of axial modulation to reduce ion-molecule reactions and can produce electron impact-like spectra that match those in the EPA/NIST Library. The mass spectrometer must be capable of producing a mass spectrum for BFB which meets all of the criteria in Table 3 when 5 or 50 ng are introduced.

4.4.4 GC/MS interface - Two alternatives are used to interface the GC to the mass spectrometer.

4.4.4.1 Direct coupling by inserting the column into the mass spectrometer is generally used for 0.25-0.32 mm id columns.

4.4.4.2 A separator including an all transfer

line and glass enrichment device or split interface is used with an 0.53 mm column.

4.4.4.3 Any enrichment device or transfer line can be used if all of the performance specifications described in Section 8 (including acceptable calibration at 50 ng or less) can be achieved. GC-to-MS interfaces constructed entirely of glass or of glass-lined materials are recommended. Glass can be deactivated by silanizing with dichlorodimethylsilane.

4.4.5 Data system - A computer system that allows the continuous acquisition and storage on machine-readable media of all mass spectra obtained throughout the duration of the chromatographic program must be interfaced to the mass spectrometer. The computer must have software that allows searching any GC/MS data file for ions of a specified mass and plotting such ion abundances versus time or scan number. This type of plot is defined as an Extracted Ion Current Profile (EICP). Software must also be available that allows integrating the abundances in any EICP between specified time or scan-number limits. The most recent version of the EPA/NIST Mass Spectral Library should also be available.

4.5 Microsyringes - 10, 25, 100, 250, 500, and 1,000 μ L.

4.6 Syringe valve - Two-way, with Luer ends (three each), if applicable to the purging device.

4.7 Syringes - 5, 10, or 25 mL, gas-tight with shutoff valve.

4.8 Balance - Analytical, 0.0001 g, and top-loading, 0.1 g.

4.9 Glass scintillation vials - 20 mL, with Teflon lined screw-caps or glass culture tubes with Teflon lined screw-caps.

4.10 Vials - 2 mL, for GC autosampler.

4.11 Disposable pipets - Pasteur.

4.12 Volumetric flasks, Class A - 10 mL and 100 mL, with ground-glass stoppers.

4.13 Spatula - Stainless steel.

5.0 REAGENTS

5.1 Reagent grade inorganic chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all inorganic reagents shall conform to the specifications of the

Committee on Analytical Reagents of the American Chemical Society, where such specifications are available. Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

5.2 Organic-free reagent water - All references to water in this method refer to organic-free reagent water, as defined in Chapter One.

5.3 Methanol, CH_3OH - Pesticide quality or equivalent, demonstrated to be free of analytes. Store apart from other solvents.

5.4 Reagent Hexadecane - Reagent hexadecane is defined as hexadecane in which interference is not observed at the method detection limit of compounds of interest.

5.4.1 In order to demonstrate that all interfering volatiles have been removed from the hexadecane, a direct injection blank must be analyzed.

5.5 Polyethylene glycol, $\text{H}(\text{OCH}_2\text{CH}_2)_n\text{OH}$ - Free of interferences at the detection limit of the target analytes.

5.6 Hydrochloric acid (1:1 v/v), HCl - Carefully add a measured volume of concentrated HCl to an equal volume of organic-free reagent water.

5.7 Stock solutions - Stock solutions may be prepared from pure standard materials or purchased as certified solutions. Prepare stock standard solutions in methanol, using assayed liquids or gases, as appropriate.

5.7.1 Place about 9.8 mL of methanol in a 10 mL tared ground-glass-stoppered volumetric flask. Allow the flask to stand, unstoppered, for about 10 minutes or until all alcohol-wetted surfaces have dried. Weigh the flask to the nearest 0.0001 g.

5.7.2 Add the assayed reference material, as described below.

5.7.2.1 Liquids - Using a 100 μL syringe, immediately add two or more drops of assayed reference material to the flask; then reweigh. The liquid must fall directly into the alcohol without contacting the neck of the flask.

5.7.2.2 Gases - To prepare standards for any compounds that boil below 30°C (e.g. bromomethane, chloroethane, chloromethane, or vinyl chloride), fill a 5 mL valved gas-tight syringe with the reference standard to the 5.0 mL mark. Lower the needle to 5 mm above the methanol meniscus. Slowly introduce the reference standard above the surface of the liquid. The heavy gas will rapidly dissolve in the methanol. Standards may also be prepared by using a lecture bottle equipped with a Hamilton Lecture Bottle Septum (#86600). Attach Teflon tubing to the side arm relief valve and direct a gentle stream of gas into the methanol meniscus.

5.7.3 Reweigh, dilute to volume, stopper, and then mix by inverting the flask several times. Calculate the concentration in milligrams per liter (mg/L) from the net gain in weight. When compound purity is assayed to be 96% or greater, the weight may be used without correction to calculate the concentration of the stock standard. Commercially prepared stock standards may be used at any concentration if they are certified by the manufacturer or by an independent source.

5.7.4 Transfer the stock standard solution into a bottle with a Teflon lined screw-cap. Store, with minimal headspace, at -10°C to -20°C and protect from light.

5.7.5 Prepare fresh standards for gases weekly or sooner if comparison with check standards indicates a problem. Reactive compounds such as 2-chloroethyl vinyl ether and styrene may need to be prepared more frequently. All other standards must be replaced after two months, or sooner if comparison with check standards indicates a problem. Both gas and liquid standards must be monitored closely by comparison to the initial calibration curve and by comparison to QC check standards. It may be necessary to replace the standards more frequently if either check exceeds a 25% drift.

5.7.6 Optionally calibration using a certified gaseous mixture can be accomplished daily utilizing commercially available gaseous analyte mixture of bromomethane, chloromethane, chloroethane, vinyl chloride, dichlorodifluoromethane and trichlorofluoromethane in nitrogen. These mixtures of documented quality are stable for as long as six months without refrigeration. (VOA-CYL III, RESTEK Corporation, Cat. #20194 or equivalent).

5.7.6.1 Preparation of Calibration Standards From
a Gas Mixture

5.7.6.1.1 Before removing the cylinder shipping cap, be sure the valve is completely closed (turn clockwise). The contents are under pressure and should be used in a well-ventilated area.

5.7.6.1.2 Wrap the pipe thread end of the Luer fitting with Teflon tape. Remove the shipping cap from the cylinder and replace it with the Luer fitting.

5.7.6.1.3 Transfer half the working standard containing other analytes, internal standards, and surrogates to the purge apparatus.

5.7.6.1.4 Purge the Luer fitting and stem on the gas cylinder prior to sample removal using the following sequence:

a) Connect either the 100 _L or 500 _L Luer syringe to the inlet fitting of the cylinder.

b) Make sure the on/off valve on the syringe is in the open position.

c) Slowly open the valve on the cylinder and withdraw a full syringe volume.

d) Be sure to close the valve on the cylinder before you withdraw the syringe from the Luer fitting.

e) Expel the gas from the syringe into a well-ventilated area.

f) Repeat steps a through e one more time to fully purge the fitting.

5.7.6.1.5 Once the fitting and stem have been purged, quickly withdraw the volume of gas you require using steps 5.6.6.1.4(a) through (d). Be sure to close the valve on the cylinder and syringe before you withdraw the syringe from the Luer fitting.

5.7.6.1.6 Open the syringe on/off valve for 5 seconds to reduce the syringe pressure to atmospheric pressure. The pressure in the cylinder is ~30psi.

5.7.6.1.7 The gas mixture should be quickly transferred into the reagent water through the female Luer fitting located above the purging vessel.

NOTE: Make sure the arrow on the 4-way valve is pointing toward the female Luer fitting when transferring the sample from the syringe. Be sure to switch the 4-way valve back to the closed position before removing the syringe from the luer fitting.

5.7.6.1.8 Transfer the remaining half of the working standard into the purging vessel. This procedure insures that the total volume of gas mix is flushed into the purging vessel, with none remaining in the valve or lines.

5.7.6.1.9 Concentration of each compound in the cylinder is typically 0.0025 g/L.

5.7.6.1.10 The following are the recommended gas volumes spiked in to 5 mLs of water to produce a typical 5-point calibration:

<u>Gas Volume</u>	<u>Calibration Concentration</u>
40 <u>L</u>	20 <u>g/L</u>
100 <u>L</u>	50 <u>g/L</u>
200 <u>L</u>	100 <u>g/L</u>
300 <u>L</u>	150 <u>g/L</u>
400 <u>L</u>	200 <u>g/L</u>

5.7.6.1.11 The following are the recommended gas volumes spiked in to 25-mls of water to produce a typical 5-point calibration:

<u>Gas Volume</u>	<u>Calibration Concentration</u>
10 <u>L</u>	1 <u>g/L</u>
20 <u>L</u>	2 <u>g/L</u>
50 <u>L</u>	5 <u>g/L</u>
100 <u>L</u>	10 <u>g/L</u>
250 <u>L</u>	25 <u>g/L</u>

5.8 Secondary dilution standards - Using stock standard solutions, prepare in methanol, secondary dilution standards containing the compounds of interest, either singly or mixed together. Secondary dilution standards must be stored with minimal headspace and should be checked frequently for signs of degradation or evaporation, especially just prior to preparing calibration standards from them. Store in a vial with no headspace for one week only.

5.9 Surrogate standards - The surrogates recommended are toluene-d₈, 4-bromofluorobenzene, and dibromofluoromethane. Other compounds may be used as surrogates, depending upon the analysis requirements. A stock surrogate solution in methanol should be prepared as described above, and a surrogate standard spiking solution should be prepared from the stock at a concentration of 50-250 g/10 mL in methanol. Each sample undergoing GC/MS analysis must be spiked with 10 L of the surrogate spiking solution prior

to analysis.

5.9.1 If a more sensitive mass spectrometer is employed to achieve lower detection levels, more dilute surrogate solutions may be required.

5.10 Internal standards - The recommended internal standards are fluorobenzene, chlorobenzene-d₅, and 1,4-dichlorobenzene-d₄. Other compounds may be used as internal standards as long as they have retention times similar to the compounds being detected by GC/MS. Prepare internal standard stock and secondary dilution standards in methanol using the procedures described in Sections 5.7 and 5.8. It is recommended that the secondary dilution standard should be prepared at a concentration of 25mg/L of each internal standard compound. Addition of 10 μL of this standard to 5.0mL of sample or calibration standard would be the equivalent of 50 μg/L.

5.10.1 If a more sensitive mass spectrometer is employed to achieve lower detection levels, more dilute internal standard solutions may be required. Area counts of the internal standard peaks should be between 50-200% of the area of the target analytes in the mid-point calibration analysis.

5.11 4-Bromofluorobenzene (BFB) standard - A standard solution containing 25 ng/_L of BFB in methanol should be prepared.

5.11.1 If a more sensitive mass spectrometer is employed to achieve lower detection levels, a more dilute BFB standard solution may be required.

5.12 Calibration standards - Calibration standards at a minimum of five concentrations should be prepared from the secondary dilution of stock standards (see Sections 5.7 and 5.8). Prepare these solutions in organic-free reagent water. One of the concentrations should be at a concentration near, but above, the method detection limit. The remaining concentrations should correspond to the expected range of concentrations found in real samples but should not exceed the working range of the GC/MS system. Each standard should contain each analyte for detection by this method (e.g. some or all of the compounds listed in Table 1 may be included). Calibration standards must be prepared daily.

5.13 Matrix spiking standards - Matrix spiking standards should be prepared from volatile organic compounds which will be representative of the compounds being investigated. At a minimum, the matrix spike should include 1,1-dichloroethene, trichloroethene, chlorobenzene, toluene, and benzene. It is desirable to perform a matrix spike using compounds found in samples. Some permits may require spiking specific compounds of interest, especially if they are polar and would not be represented by the above listed compounds. The standard should be prepared in methanol, with each compound present at a concentration of 250 _g/10.0 mL.

5.13.1 If a more sensitive mass spectrometer is employed to achieve lower detection levels, more dilute matrix spiking solutions may be required.

5.14 Great care must be taken to maintain the integrity of all standard solutions. It is recommended all standards in methanol be stored at -10°C to -20°C in amber bottles with Teflon lined screw-caps.

6.0 SAMPLE COLLECTION, PRESERVATION, AND HANDLING

6.1 See the introductory material to this chapter, Organic Analytes, Section 4.1.

7.0 PROCEDURE

7.1 Three alternate methods are provided for sample introduction. All internal standards, surrogates, and matrix spikes (when applicable) must be added to samples before introduction.

7.1.1 Direct injection - in very limited application, (e.g., volatiles in waste oil or aqueous process wastes) direct injection of aqueous samples or samples diluted according to Method 3585 may be appropriate. Direct injection has been used for the analysis of volatiles in waste oil (diluted 1:1 with hexadecane) and for determining if the sample is ignitable (aqueous injection, Methods 1010 or 1020). Direct injection is only permitted for the determination of volatiles at the TCLP regulatory limits, at concentrations in excess of 10,000_g/L, or for water-soluble compounds that do not purge.

7.1.2 Purge-and-trap for aqueous samples, see Method 5030 for details.

7.1.3 Purge-and-trap for solid samples, see Method 5030~~5~~ for details.

7.2 Chromatographic conditions

7.2.1 General:

Injector temperature: 200-225°C
Transfer line temperature: 250-300°C

7.2.2 Column 1 (A sample chromatogram is presented in Figure 5)

Carrier gas (He) flow rate: 15 mL/min
Initial temperature: 10°C, hold for 5 minutes
Temperature program: 6°C/min to 160°C
Final temperature: 160°C, hold until all expected compounds have eluted.

7.2.3 Column 2, Cryogenic cooling (A sample chromatogram is presented in Figure 6)

Carrier gas (He) flow rate: 15 mL/min

Initial temperature: 10°C, hold for 5 minutes
Temperature program: 6°C/min to 160°C
Final temperature: 160°C, hold until all expected compounds have eluted.

7.2.4 Column 2, Non-cryogenic cooling (A sample chromatogram is presented in Figure 7). It is recommended that carrier gas flow and split and make-up gases be set using performance of standards as guidance. Set the carrier gas head pressure to 10 psi and the split to 30 mL/min. Optimize the make-up gas flow for the separator (approximately 30 mL/min) by injecting BFB, and determining the optimum response when varying the make-up gas. This will require several injections of BFB. Next, make several injections of the volatile working standard with all analytes of interest. Adjust the carrier and split to provide optimum chromatography and response. This is an especially critical adjustment for the volatile gas analytes. The head pressure should optimize between 8-12 psi and the split between 20-60 mL/min. The use of the splitter is important to minimize the effect of water on analyte response, to allow the use of a larger volume of helium during trap desorption, and to slow column flow.

Initial temperature: 45°C, hold for 2 minutes
Temperature program: 8°C/min to 200°C
Final temperature: 200°C, hold for 6 minutes.

A trap preheated to 150°C prior to trap desorption is required to provide adequate chromatography of the gas analytes.

7.2.5 Column 3 (A sample chromatogram is presented in Figure 8)

Carrier gas (He) flow rate: 4 mL/min
Initial temperature: 10°C, hold for 5 minutes
Temperature program: 6°C/min to 70°C, then 15°C/min to 145°C
Final temperature: 145°C, hold until all expected compounds have eluted.

7.2.6 Direct injection - Column 4

Carrier gas (He) flow rate: 4 mL/min
Column: J&W DB-24, 70m x 0.53 mm
Initial temperature: 40°C, hold for 3 minutes
Temperature program: 8°C/min
Final temperature: 260°C, hold until all

expected compounds have eluted.
Column Bake out (direct inj): 75 minutes
Injector temperature: 200-225_C
Transfer line temperature: 250-300_C

7.3 Initial calibration - the recommended MS operating conditions

Mass range: 35-260 amu
Scan time: 0.6-2 sec/scan
Source temperature: According to manufacturer's specifications
Ion trap only: Axial modulation 4.0 volts
Manifold set 220_C
Emission current 30_amps

7.3.1 Each GC/MS system must be hardware-tuned to meet the criteria in Table 4 for a 5-50 ng injection or purging of 4-bromofluorobenzene (2 μ L injection of the BFB standard). Analyses must not begin until these criteria are met.

7.3.2 **Set-up the purge and trap system as outlined in Method 5030 if purge and trap analysis is to be utilized.** A set of at least five calibration standards containing the method analytes is needed. One calibration standard should contain each analyte at a concentration approaching but greater than the method detection limit (Table 1) for that compound; the other calibration standards should contain analytes at concentrations that define the range of the method. Calibration should be done using the sample introduction technique that will be used for samples. For Method 5030, the purging efficiency for 5 mL of water is greater than for 25 mL. Therefore, develop the standard curve with whichever volume of sample that will be analyzed.

7.3.2.1 To prepare a calibration standard for purge-and-trap or aqueous direct injection, add an appropriate volume of a secondary dilution standard solution to an aliquot of organic-free reagent water in a volumetric flask. Use a microsyringe and rapidly inject the alcoholic standard into the expanded area of the filled volumetric flask. Remove the needle as quickly as possible after injection. Mix by inverting the flask three times only. Discard the contents contained in the neck of the flask. Aqueous standards are not stable and should be prepared daily. Transfer 5.0 mL (or 25 mL if lower detection limits are required) of each standard to a gas tight syringe along with 10 μ L of internal standard. Then transfer the contents to a purging device or syringe. **Perform purge and trap or direct injection as outlined in Method 5030.**

7.3.2.2 To prepare a calibration standard for direct injection analysis of oil, dilute standards in hexadecane.

7.3.3 Tabulate the area response of the characteristic ions (see Table 5) against concentration for each compound and each internal standard. Calculate response factors (RF) for each compound relative to one of the internal standards. The internal standard selected for the calculation of the RF for a compound should be the internal standard that has a retention time closest to the compound being measured (Section 7.6.2). The RF is calculated as follows:

$$RF = (A_x C_{is}) / (A_{is} C_x)$$

where:

A_x = Area of the characteristic ion for the compound being measured.
 A_{is} = Area of the characteristic ion for the specific internal standard.
 C_{is} = Concentration of the specific internal standard.
 C_x = Concentration of the compound being measured.

7.3.4 The average RRF must be calculated and recorded for each compound. A system performance check should be made before this calibration curve is used. Five compounds (the System Performance Check Compounds, or SPCCs) are checked for a minimum average relative response factor. These compounds are chloromethane; 1,1-dichloroethane; bromoform; 1,1,2,2-tetrachloroethane; and chlorobenzene. ~~The minimum acceptable average RRF for these compounds should be 0.100.~~

These compounds are used to check compound instability and to check for degradation caused by contaminated lines or active sites in the system. Examples of these occurrences are:

7.3.4.1 Chloromethane - This compound is the most likely compound to be lost if the purge flow is too fast.

7.3.4.2 Bromoform - This compound is one of the compounds most likely to be purged very poorly if the purge flow is too slow. Cold spots and/or active sites in the transfer lines may adversely affect response. Response of the quantitation ion (m/z 173) is

directly affected by the tuning of BFB at ions m/z 174/176. Increasing the m/z 174/176 ratio relative to m/z 95 may improve bromoform response.

7.3.4.3 Tetrachloroethane and 1,1-dichloroethane - These compounds are degraded by contaminated transfer lines in purge-and-trap systems and/or active sites in trapping materials.

7.3.5 Using the RRFs from the initial calibration, calculate and record the percent relative standard deviation (%RSD) for all compounds. The percent RSD is calculated as follows:

where:

RSD = Relative standard deviation.
RF_x = mean of 5 initial RRFs for a compound.
SD = standard deviation of average RRFs for a compound.

$$SD = \frac{\sqrt{\sum_{i=1}^N (x_i - \bar{x})^2}}{N - 1}$$

The percent relative standard deviation should be less than 15% for each compound. However, the %RSD for each individual Calibration Check Compound (CCC) must be less than 30%. The CCCs are:

1,1-Dichloroethene,
Chloroform,
1,2-Dichloropropane,
Toluene,
Ethylbenzene, and
Vinyl chloride.

7.3.5.1 If a %RSD greater than 30 percent is measured for any CCC, then corrective action to eliminate a system leak and/or column reactive sites is required before reattempting calibration.

7.3.6 Linearity - If the %RSD of any compound is 15% or less, then the relative response factor is assumed to be constant over the calibration range, and the average relative response factor may be used for quantitation.

7.3.6.1 If the %RSD of any compound is greater

than 15%, construct calibration curves of area ratio (A/A_{is}) versus concentration using first or higher order regression fit of the five calibration points. The analyst should select the regression order which introduces the least calibration error into the quantitation. If the %RSD is <15%, use of calibration curves is a recommended alternative to average response factor calibration, and a useful diagnostic of standard preparation accuracy and absorption activity in the chromatographic system.

7.3.7 These curves are verified each shift by purging a performance standard. Recalibration is required only if calibration and on-going performance criteria cannot be met.

7.4 Daily GC/MS calibration verification

7.4.1 Prior to the analysis of samples, inject or purge 5-50 ng of the 4-bromofluorobenzene standard following Method 5030. The resultant mass spectra for the BFB must meet all of the criteria given in Table 4 before sample analysis begins. These criteria must be demonstrated each 12-hour shift.

7.4.2 The initial calibration curve (Section 7.3) for each compound of interest must be checked and verified once every 12 hours during analysis with the introduction technique used for samples. This is accomplished by analyzing a calibration standard that is at a concentration near the midpoint concentration for the working range of the GC/MS by checking the SPCC and CCC.

7.4.3 System Performance Check Compounds (SPCCs) - A system performance check must be made each 12 hours. If the SPCC criteria are met, a comparison of relative response factors is made for all compounds. This is the same check that is applied during the initial calibration. If the minimum relative response factors are not met, the system must be evaluated, and corrective action must be taken before sample analysis begins. Some possible problems are standard mixture degradation, injection port inlet contamination, contamination at the front end of the analytical column, and active sites in the column or chromatographic system.

7.4.3.1 The minimum relative response factor for volatile SPCCs are as follows:

Chloromethane	0.1
1,1-Dichloroethane	0.1
Bromoform	0.25

Chlorobenzene
1,1,2,2-Tetrachloroethane

0.3

0.3

7.4.4 Calibration Check Compounds (CCCs) - After the system performance check is met, CCCs listed in Section 7.3.8 are used to check the validity of the initial calibration.

Calculate the percent drift using the following equation:

$$\% \text{ Drift} = (C_i - C_c) / C_i \times 100$$

where:

C_i = Calibration Check Compound standard
concentration.
 C_c = Measured concentration using selected
quantitation method.

If the percent drift for each CCC is less than 20%, the initial calibration is assumed to be valid. If the criterion is not met (> 20% drift), for any one CCC, corrective action must be taken. Problems similar to those listed under SPCCs could affect this criterion. If no source of the problem can be determined after corrective action has been taken, a new five point calibration MUST be generated. This criterion MUST be met before quantitative sample analysis begins. If the CCCs are not required analytes by the permit, then all required analytes must meet the 20% drift criterion.

7.4.5 The internal standard responses and retention times in the check calibration standard must be evaluated immediately after or during data acquisition. If the retention time for any internal standard changes by more than 30 seconds from the last check calibration (12 hours), the chromatographic system must be inspected for malfunctions and corrections must be made, as required. If the EICP area for any of the internal standards changes by a factor of two (-50% to +100%) from the last daily calibration standard check, the mass spectrometer must be inspected for malfunctions and corrections must be made, as appropriate. When corrections are made, reanalysis of samples analyzed while the system was malfunctioning are necessary.

7.4.6 For compounds which exhibit linearity of response, the RRF of the daily check standard may be used for quantitation, provided the criteria for SPCCs and CCCs are satisfied.

7.5 GC/MS analysis

7.5.1 It is highly recommended that the extract be screened on a headspace-GC/FID (Methods 3810/8015), headspace-GC/PID/ELCD (Methods 3810/8021), or waste dilution-GC/PID/ELCD (Methods 3585/8021) using the same type of capillary column. This will minimize contamination of the GC/MS system from unexpectedly high concentrations of organic compounds. Use of screening is particularly important when this method is used to achieve low detection levels.

7.5.2 All samples and standard solutions must be allowed to warm to ambient temperature before analysis.

Set-up the purge and trap system as outlined in Method 5030 if purge and trap introduction will be used.

7.5.3 BFB tuning criteria and daily GC/MS calibration verification criteria must be met before analyzing samples.

7.5.3.1 Remove the plunger from a 5 mL syringe and attach a closed syringe valve. If lower detection limits are required, use a 25 mL syringe. Open the sample or standard bottle, which has been allowed to come to ambient temperature, and carefully pour the sample

into the syringe barrel to just short of overflowing. Replace the syringe plunger and compress the sample. Open the syringe valve and vent any residual air while adjusting the sample volume to 5.0 mL.

7.5.4 The process of taking an aliquot destroys the validity of aqueous and soil samples for future analysis; therefore, if there is only one VOA vial, the analyst should prepare a second aliquot for analysis at this time to protect against possible loss of sample integrity. This second sample is maintained only until such time when the analyst has determined that the first sample has been analyzed properly. For aqueous samples, filling one 20 mL syringe would require the use of only one syringe. If a second analysis is needed from a syringe, it must be analyzed within 24 hours. Care must be taken to prevent air from leaking into the syringe.

7.5.4.1 The following procedure is appropriate for diluting aqueous purgeable samples. All steps must be performed without delays until the diluted sample is in a gas-tight syringe.

7.5.4.1.1 Dilutions may be made in volumetric flasks (10 to 100 mL). Select the volumetric flask that will allow for the necessary dilution. Intermediate dilutions may be necessary for extremely large dilutions.

7.5.4.1.2 Calculate the approximate volume of organic-free reagent water to be added to the volumetric flask selected and add slightly less than this quantity of organic-free reagent water to the flask.

7.5.4.1.3 Inject the proper aliquot of sample from the syringe into the flask. Aliquots of less than 1 mL are not recommended. Dilute the sample to the mark with organic-free reagent water. Cap the flask, invert, and shake three times. Repeat above procedure for additional dilutions.

7.5.4.1.4 Fill a 5 mL syringe with the diluted sample.

7.5.4.2 Compositing aqueous samples prior to GC/MS analysis

7.5.4.2.1 Add 5 mL or equal larger amounts of each sample (up to 5 samples are allowed) to a 25 mL glass syringe. Special precautions must

be made to maintain zero headspace in the syringe.

7.5.4.2.2 The samples must be cooled at 4°C during this step to minimize volatilization losses.

7.5.4.2.3 Mix well and draw out a 5 mL aliquot for analysis.

7.5.4.2.4 Follow sample introduction, purging, and desorption steps described in Method 5030.

7.5.4.2.5 If less than five samples are used for compositing, a proportionately smaller syringe may be used unless a 25 mL sample is to be purged.

7.5.5 Add 10.0 μ L of surrogate spiking solution and 10 μ L of internal standard spiking solution to each sample. The surrogate and internal standards may be mixed and added as a single spiking solution. The addition of 10 μ L of the surrogate spiking solution to 5 mL of sample is equivalent to a concentration of 50 μ g/L of each surrogate standard. The addition of 10 μ L of the surrogate spiking solution to 5 g of sample is equivalent to a concentration of 50 μ g/kg of each surrogate standard.

7.5.5.1 If a more sensitive mass spectrometer is employed to achieve lower detection levels, more dilute surrogate and internal standard solutions may be required.

7.5.6 Perform purge and trap or direct injection by Method 5030. If the initial analysis of sample or a dilution of the sample has a concentration of analytes that exceeds the initial calibration range, the sample must be reanalyzed at a higher dilution. Secondary ion quantitation is allowed only when there are sample interferences with the primary ion. When a sample is analyzed that has saturated ions from a compound, this analysis must be followed by a blank organic-free reagent water analysis. If the blank analysis is not free of interferences, the system must be decontaminated. Sample analysis may not resume until the blank analysis is demonstrated to be free of interferences.

7.5.6.1. All dilutions should keep the response of the major constituents (previously saturated peaks) in the upper half of the linear range of the curve. Proceed to Sections 7.6.1 and 7.6.2 for qualitative and

quantitative analysis.

7.5.7 For matrix spike analysis, add 10 μ L of the matrix spike solution (Section 5.13) to the 5 mL of sample to be purged. Disregarding any dilutions, this is equivalent to a concentration of 50 μ g/L of each matrix spike standard.

7.6 Data interpretation

7.6.1 Qualitative analysis

7.6.1.1 The qualitative identification of compounds determined by this method is based on retention time, and on comparison of the sample mass spectrum, after background correction, with characteristic ions in a reference mass spectrum. The reference mass spectrum must be generated by the laboratory using the conditions of this method. The characteristic ions from the reference mass spectrum are defined to be the three ions of greatest relative intensity, or any ions over 30% relative intensity if less than three such ions occur in the reference spectrum. Compounds should be identified as present when the criteria below are met.

7.6.1.1.1 The intensities of the characteristic ions of a compound maximize in the same scan or within one scan of each other. Selection of a peak by a data system target compound search routine where the search is based on the presence of a target chromatographic peak containing ions specific for the target compound at a compound-specific retention time will be accepted as meeting this criterion.

7.6.1.1.2 The RRT of the sample component is within \pm 0.06 RRT units of the RRT of the standard component.

7.6.1.1.3 The relative intensities of the characteristic ions agree within 30% of the relative intensities of these ions in the reference spectrum. (Example: For an ion with an abundance of 50% in the reference spectrum, the corresponding abundance in a sample spectrum can range between 20% and 80%.)

7.6.1.1.4 Structural isomers that produce very similar mass spectra should be identified as individual isomers if they have sufficiently

different GC retention times. Sufficient GC resolution is achieved if the height of the valley between two isomer peaks is less than 25% of the sum of the two peak heights. Otherwise, structural isomers are identified as isomeric pairs.

7.6.1.1.5 Identification is hampered when sample components are not resolved chromatographically and produce mass spectra containing ions contributed by more than one analyte. When gas chromatographic peaks obviously represent more than one sample component (i.e., a broadened peak with shoulder(s) or a valley between two or more maxima), appropriate selection of analyte spectra and background spectra is important. Examination of extracted ion current profiles of appropriate ions can aid in the selection of spectra, and in qualitative identification of compounds. When analytes coelute (i.e., only one chromatographic peak is apparent), the identification criteria can be met, but each analyte spectrum will contain extraneous ions contributed by the coeluting compound.

7.6.1.2 For samples containing components not associated with the calibration standards, a library search may be made for the purpose of tentative identification. The necessity to perform this type of identification will be determined by the type of analyses being conducted. Guidelines for making tentative identification are:

(1) Relative intensities of major ions in the reference spectrum (ions > 10% of the most abundant ion) should be present in the sample spectrum.

(2) The relative intensities of the major ions should agree within $\pm 20\%$. (Example: For an ion with an abundance of 50% in the standard spectrum, the corresponding sample ion abundance must be between 30 and 70%).

(3) Molecular ions present in the reference spectrum should be present in the sample spectrum.

(4) Ions present in the sample spectrum but not in the reference spectrum should be reviewed for possible background contamination or presence of coeluting compounds.

(5) Ions present in the reference spectrum but not in the sample spectrum should be reviewed for possible subtraction from the sample spectrum because of background contamination or coeluting peaks. Data system library reduction programs can sometimes create these discrepancies.

Computer generated library search routines should not use normalization routines that would misrepresent the library or unknown spectra when compared to each other. Only after visual comparison of sample with the nearest library searches will the mass spectral interpretation specialist assign a tentative identification.

7.6.2 Quantitative analysis

7.6.2.1 When a compound has been identified, the quantitation of that compound will be based on the integrated abundance from the EICP of the primary characteristic ion. Quantitation will take place using the internal standard technique. The internal standard used shall be the one nearest the retention time of that of a given analyte.

7.6.2.2 When MS response is linear and passes through the origin, calculate the concentration of each identified analyte in the sample as follows:

Water

$$\text{concentration (g/L)} = \frac{(A_x)(I_s)}{(A_{is})(\text{RRF})(V_o)}$$

where:

A_x = Area of characteristic ion for compound being measured.
 I_s = Amount of internal standard injected (ng).
 A_{is} = Area of characteristic ion for the internal standard.
 RRF = Relative Response factor for compound being measured.
 V_o = Volume of water purged (mL), taking into consideration any dilutions made.

Sediment/Soil Sludge (on a dry-weight basis) and
Waste (normally on a wet-weight basis)

$$\text{concentration (g/kg)} = \frac{(A_x) (I_s) (V_t)}{(A_{is}) (\text{RRF}) (V_i) (W_s) (D)}$$

where:

$A_x, I_s, A_{is}, \text{RRF}$, = Same as for water.
 V_t = Volume of total extract (_L)
(use 10,000 _L or a factor of this when
dilutions are made).
 V_i = Volume of extract added (_L) for purging.
 W_s = Weight of sample extracted or purged (g).
 D = % dry weight of sample/100, or 1 for a
wet-weight basis.

7.6.2.3 Where applicable, an estimate of concentration for noncalibrated components in the sample should be made. The formulae given above should be used with the following modifications: The areas A_x and A_{is} should be from the total ion chromatograms, and the RRF for the compound should be assumed to be 1. The concentration obtained should be reported indicating (1) that the value is an estimate and (2) which internal standard was used to determine concentration. Use the nearest internal standard free of interferences.

8.0 QUALITY CONTROL

8.1 Refer to Chapter One and Method 8000 for general quality control procedures.

8.2 Additional required instrument QC is found in the Sections 7.3 and 7.4:

8.2.1 The GC/MS system must be tuned to meet the BFB specifications.

8.2.2 There must be an initial calibration of the GC/MS system

8.2.3 The GC/MS system must meet the SPCC criteria and the CCC criteria, each 12 hours.

8.3 To establish the ability to generate acceptable accuracy

and precision, the analyst must perform the following operations.

8.3.1 A quality control (QC) reference sample concentrate is required containing each analyte at a concentration of 10 mg/L or less in methanol. The QC reference sample concentrate may be prepared from pure standard materials or purchased as certified solutions. If prepared by the laboratory, the QC reference sample concentrate must be made using stock standards prepared independently from those used for calibration.

8.3.2 Prepare a QC reference sample to contain 20 μ g/L or less of each analyte by adding 200 μ L of QC reference sample concentrate to 100mL of organic-free reagent water.

8.3.3 Four 5 mL aliquots of the well mixed QC reference sample are analyzed according to the method beginning in Section 7.5.1.

8.3.4 Calculate the average recovery (\bar{x}) in μ g/L, and the standard deviation of the recovery (s) in μ g/L, for each analyte using the four results.

8.3.5 Tables 7 and 8 provide single laboratory recovery and precision data obtained for the method analytes from water. Similar results from dosed water should be expected by any experienced laboratory. Compare s and \bar{x} (Section 8.3.4) for each analyte to the single laboratory recovery and precision data. Results are comparable if the calculated standard deviation of the recovery does not exceed 2.6 times the single laboratory RSD or 20%, whichever is greater, and the mean recovery lies within the interval $\bar{x} \pm 3S$ or $\bar{x} \pm 30\%$, whichever is greater.

NOTE: The large number of analytes in Tables 7 and 8 present a substantial probability that one or more will fail at least one of the acceptance criteria when all analytes of a given method are determined.

8.3.6 When one or more of the analytes tested are not comparable to the data in Table 6 or 7, the analyst must proceed according to Section 8.3.6.1 or 8.3.6.2.

8.3.6.1 Locate and correct the source of the problem and repeat the test for all analytes beginning with Section 8.3.2.

8.3.6.2 Beginning with Section 8.3.2, repeat the test only for those analytes that are not comparable.

Repeated failure, however, will confirm a general problem with the measurement system. If this occurs, locate and correct the source of the problem and repeat the test for all compounds of interest beginning with Section 8.3.2.

8.4 For aqueous and soil matrices, laboratory established surrogate control limits should be compared with the control limits listed in Table 8.

8.4.1 If recovery is not within limits, the following procedures are required.

8.4.1.1 Check to be sure that there are no errors in the calculations, surrogate solutions or internal standards. If errors are found, recalculate the data accordingly.

8.4.1.2 Check instrument performance. If an instrument performance problem is identified, correct the problem and re-analyze the extract.

8.4.1.3 If no problem is found, re-extract and re-analyze the sample.

8.4.1.4 If, upon re-analysis, the recovery is again not within limits, flag the data as "estimated concentration".

8.4.2 At a minimum, each laboratory should update surrogate recovery limits on a matrix-by-matrix basis, annually.

9.0 METHOD PERFORMANCE

9.1 The method detection limit (MDL) is defined as the minimum concentration of a substance that can be measured and reported with 99% confidence that the value is above zero. The MDL actually achieved in a given analysis will vary depending on instrument sensitivity and matrix effects.

9.2 This method has been tested in a single laboratory using spiked water. Using a wide-bore capillary column, water was spiked at concentrations between 0.5 and 10 $\mu\text{g/L}$. Single laboratory accuracy and precision data are presented for the method analytes in Table 6. Calculated MDLs are presented in Table 1.

9.3 The method was tested using water spiked at 0.1 to 0.5 $\mu\text{g/L}$ and analyzed on a cryofocussed narrow-bore column. The accuracy and precision data for these compounds are presented in Table 7. MDL values were also calculated from these data and are presented in Table 2.

9.4 Direct injection has been used for the analysis of waste motor oil samples using a wide-bore column. The accuracy and precision data for these compounds are presented in Table 10.

10.0 REFERENCES

1. Methods for the Determination of Organic Compounds in Finished Drinking Water and Raw Source Water Method 524.2; U.S. Environmental Protection Agency. Office of Research Development. Environmental Monitoring and Support Laboratory: Cincinnati, OH 1986.
2. U.S. EPA Contract Laboratory Program, Statement of Work for Organic Analysis, July 1985, Revision.
3. Bellar, T.A.; J.J. Lichtenberg. J. Amer. Water Works Assoc. 1974, 66(12), 739-744.

4. Bellar, T.A.; J.J. Lichtenberg. "Semi-Automated Headspace Analysis of Drinking Waters and Industrial Waters for Purgeable Volatile Organic Compounds"; in Van Hall, Ed.; Measurement of Organic Pollutants in Water and Wastewater, ASTM STP 686, pp 108-129, 1979.

5. Budde, W.L.; J.W. Eichelberger. "Performance Tests for the Evaluation of Computerized Gas Chromatography/Mass Spectrometry Equipment and Laboratories"; U.S. Environmental Protection Agency. Environmental Monitoring and Support Laboratory. Cincinnati, OH 45268, April 1980; EPA-600/4-79-020.
6. Eichelberger, J.W.; L.E. Harris; W.L. Budde. "Reference Compound to Calibrate Ion Abundance Measurement in Gas Chromatography-Mass Spectrometry Systems"; Analytical Chemistry 1975, 47, 995-1000.
7. Olynyk, P.; W.L. Budde; J.W. Eichelberger. "Method Detection Limit for Methods 624 and 625"; Unpublished report, October 1980.
8. Non Cryogenic Temperatures Program and Chromatogram, Private Communications; Myron Stephenson and Frank Allen, EPA Region IV Laboratory, Athens, GA.
9. Marsden, P.; C.L. Helms, B.N. Colby. "Analysis of Volatiles in Waste Oil"; report for B. Lesnik OSW/EPA under SAIC contract 68-W9-001, 6/92.
10. Methods for the Determination of Organic Compounds in Drinking Water, Supplement II Method 524.2; U.S. Environmental Protection Agency. Office of Research and Development. Environmental Monitoring Systems Laboratory: Cincinnati, OH 1992.

TABLE 1.
 CHROMATOGRAPHIC RETENTION TIMES AND METHOD DETECTION LIMITS (MDL)
 FOR VOLATILE ORGANIC COMPOUNDS ON WIDE BORE CAPILLARY COLUMNS

ANALYTE MDL ^d (_g/L)	RETENTION TIME (minutes)			
	Column 1 ^a	Column 2 ^b	Column 2' ^c	
	Dichlorodifluoromethane	1.35	0.70	3.13
Chloromethane	1.49	0.73	3.40	0.13
Vinyl Chloride	1.56	0.79	3.93	0.17
Bromomethane	2.19	0.96	4.80	0.11
Chloroethane	2.21	1.02	--	0.10
Trichlorofluoromethane	2.42	1.19	6.20	0.08
Acrolein	3.19			
Iodomethane	3.56			
Acetonitrile	4.11			
Carbon disulfide	4.11			
Allyl chloride	4.11			
Methylene chloride	4.40	2.06	9.27	0.03
1,1-Dichloroethene	4.57	1.57	7.83	0.12
Acetone	4.57			
trans-1,2-Dichloroethene	4.57	2.36	9.90	0.06
Acrylonitrile	5.00			
1,1-Dichloroethane	6.14	2.93	10.80	0.04
Vinyl acetate	6.43			
2,2-Dichloropropane	8.10	3.80	11.87	0.35
2-Butanone	--			
cis-1,2-Dichloroethene	8.25	3.90	11.93	0.12
Propionitrile	8.51			
Chloroform	9.01	4.80	12.60	0.03
Bromochloromethane	--	4.38	12.37	0.04
Methacrylonitrile	9.19			
1,1,1-Trichloroethane	10.18	4.84	12.83	0.08
Carbon tetrachloride	11.02	5.26	13.17	0.21
1,1-Dichloropropene	--	5.29	13.10	0.10
Benzene	11.50	5.67	13.50	0.04
1,2-Dichloroethane	12.09	5.83	13.63	0.06
Trichloroethene	14.03	7.27	14.80	0.19
1,2-Dichloropropane	14.51	7.66	15.20	0.04
Bromodichloromethane	15.39	8.49	15.80	0.08
Dibromomethane	15.43	7.93	15.43	0.24
Methyl methacrylate	15.50			

1,4-Dioxane	16.17			
2-Chloroethyl vinyl ether	--			
4-Methyl-2-pentanone	17.32			
trans-1,3-Dichloropropene	17.47	--	16.70	--
Toluene	18.29	10.00	17.40	0.11
cis-1,3-Dichloropropene	19.38	--	17.90	--
1,1,2-Trichloroethane	19.59	11.05	18.30	0.10
Ethyl methacrylate	20.01			

TABLE 1.
(Continued)

ANALYTE MDL ^d — (_g/L)	RETENTION TIME (minutes)			
	Column 1 ^a	Column 2 ^b	Column 2 ^c	
2-Hexanone	20.30			
Tetrachloroethene	20.26	11.15	18.60	0.14
1,3-Dichloropropane	20.51	11.31	18.70	0.04
Dibromochloromethane	21.19	11.85	19.20	0.05
1,2-Dibromoethane	21.52	11.83	19.40	0.06
1-Chlorohexane	--	13.29	--	0.05
Chlorobenzene	23.17	13.01	20.67	0.04
1,1,1,2-Tetrachloroethane	23.36	13.33	20.87	0.05
Ethylbenzene	23.38	13.39	21.00	0.06
p-Xylene	23.54	13.69	21.30	0.13
m-Xylene	23.54	13.68	21.37	0.05
o-Xylene	25.16	14.52	22.27	0.11
Styrene	25.30	14.60	22.40	0.04
Bromoform	26.23	14.88	22.77	0.12
Isopropylbenzene (Cumene)	26.37	15.46	23.30	0.15
cis-1,4-Dichloro-2-butene	27.12			
1,1,2,2-Tetrachloroethane	27.29	16.35	24.07	0.04
Bromobenzene	27.46	15.86	24.00	0.03
1,2,3-Trichloropropane	27.55	16.23	24.13	0.32
n-Propylbenzene	27.58	16.41	24.33	0.04
2-Chlorotoluene	28.19	16.42	24.53	0.04
trans-1,4-Dichloro-2-butene	28.26			
1,3,5-Trimethylbenzene	28.31	16.90	24.83	0.05
4-Chlorotoluene	28.33	16.72	24.77	0.06
Pentachloroethane	29.41			
1,2,4-Trimethylbenzene	29.47	17.70	31.50	0.13
sec-Butylbenzene	30.25	18.09	26.13	0.13
tert-Butylbenzene	30.59	17.57	26.60	0.14

p-Isopropyltoluene	30.59	18.52	26.50	0.12
1,3-Dichlorobenzene	30.56	18.14	26.37	0.12
1,4-Dichlorobenzene	31.22	18.39	26.60	0.03
Benzyl chloride	32.00			
n-Butylbenzene	32.23	19.49	27.32	0.11
1,2-Dichlorobenzene	32.31	19.17	27.43	0.03
1,2-Dibromo-3-chloropropane	35.30	21.08	--	0.26
1,2,4-Trichlorobenzene	38.19	23.08	31.50	0.04
Hexachlorobutadiene	38.57	23.68	32.07	0.11
Naphthalene	39.05	23.52	32.20	0.04
1,2,3-Trichlorobenzene	40.01	24.18	32.97	0.03

TABLE 1.
(Continued)

ANALYTE MDL ^d (<u>g/L</u>)	RETENTION TIME (minutes)		
	Column 1 ^a	Column 2 ^b	Column 2' ^c
INTERNAL STANDARDS/SURROGATES			
1,4-Difluorobenzene	13.26		
Chlorobenzene-d ₅	23.10		
1,4-Dichlorobenzene-d ₄	31.16		
4-Bromofluorobenzene	27.83	15.71	23.63
1,2-Dichlorobenzene-d ₄	32.30	19.08	27.25
Dichloroethane-d ₄	12.08		
Dibromofluoromethane	--		
Toluene-d ₈	18.27		
Pentafluorobenzene	--		
Fluorobenzene	13.00	6.27	14.06

^a Column 1 - 60 meter x 0.75 mm ID VOCOL capillary. Hold at 10°C for 8 minutes, then program to 180°C at 4°/min.

^b Column 2 - 30 meter x 0.53 mm ID DB-624 wide-bore capillary using cryogenic oven. Hold at 10°C for 5 minutes, then program to 160°C at 6°/min.

^c Column 2' - 30 meter x 0.53 mm ID DB-624 wide-bore capillary, cooling GC oven to ambient temperatures. Hold at 10°C for 6 minutes, program to 70°C at 10°/min, program to 120°C at 5°/min, then program to 180°C at 8°/min.

^d MDL based on a 25 mL sample volume.

TABLE 2.
 CHROMATOGRAPHIC RETENTION TIMES AND METHOD DETECTION LIMITS (MDL)
 FOR VOLATILE ORGANIC COMPOUNDS ON NARROW BORE CAPILLARY COLUMNS

ANALYTE	RETENTION TIME (minutes) Column 3 ^a	MDL ^b (<u>g/L</u>)
Dichlorodifluoromethane	0.88	0.11
Chloromethane	0.97	0.05
Vinyl chloride	1.04	0.04
Bromomethane	1.29	0.06
Chloroethane	1.45	0.02
Trichlorofluoromethane	1.77	0.07
1,1-Dichloroethene	2.33	0.05
Methylene chloride	2.66	0.09
trans-1,2-Dichloroethene	3.54	0.03
1,1-Dichloroethane	4.03	0.03
cis-1,2-Dichloroethene	5.07	0.06
2,2-Dichloropropane	5.31	0.08
Chloroform	5.55	0.04
Bromochloromethane	5.63	0.09
1,1,1-Trichloroethane	6.76	0.04
1,2-Dichloroethane	7.00	0.02
1,1-Dichloropropene	7.16	0.12
Carbon tetrachloride	7.41	0.02
Benzene	7.41	0.03
1,2-Dichloropropane	8.94	0.02
Trichloroethene	9.02	0.02
Dibromomethane	9.09	0.01
Bromodichloromethane	9.34	0.03
Toluene	11.51	0.08
1,1,2-Trichloroethane	11.99	0.08
1,3-Dichloropropane	12.48	0.08
Dibromochloromethane	12.80	0.07
Tetrachloroethene	13.20	0.05
1,2-Dibromoethane	13.60	0.10
Chlorobenzene	14.33	0.03
1,1,1,2-Tetrachloroethane	14.73	0.07
Ethylbenzene	14.73	0.03
p-Xylene	15.30	0.06
m-Xylene	15.30	0.03
Bromoform	15.70	0.20
o-Xylene	15.78	0.06
Styrene	15.78	0.27
1,1,2,2-Tetrachloroethane	15.78	0.20

1,2,3-Trichloropropane
Isopropylbenzene

16.26
16.42

0.09
0.10

TABLE 2.
(Continued)

ANALYTE	RETENTION TIME (minutes) Column 3 ^a	MDL ^b (<u>g/L</u>)
Bromobenzene	16.42	0.11
2-Chlorotoluene	16.74	0.08
n-Propylbenzene	16.82	0.10
4-Chlorotoluene	16.82	0.06
1,3,5-Trimethylbenzene	16.99	0.06
tert-Butylbenzene	17.31	0.33
1,2,4-Trimethylbenzene	17.31	0.09
sec-Butylbenzene	17.47	0.12
1,3-Dichlorobenzene	17.47	0.05
p-Isopropyltoluene	17.63	0.26
1,4-Dichlorobenzene	17.63	0.04
1,2-Dichlorobenzene	17.79	0.05
n-Butylbenzene	17.95	0.10
1,2-Dibromo-3-chloropropane	18.03	0.50
1,2,4-Trichlorobenzene	18.84	0.20
Naphthalene	19.07	0.10
Hexachlorobutadiene	19.24	0.10
1,2,3-Trichlorobenzene	19.24	0.14

^a Column 3 - 30 meter x 0.32 mm ID DB-5 capillary with 1 m film thickness.

^b MDL based on a 25 mL sample volume.

TABLE 3.
ESTIMATED QUANTITATION LIMITS FOR VOLATILE ANALYTES^a

	Estimated Quantitation Limits	
	Ground water _g/L	Low Soil/Sediment ^b _g/kg
Volume of water purged	5 mL	25 mL
All analytes in Table 1	5	1
		5

^a Estimated Quantitation Limit (EQL) - The lowest concentration that can be reliably achieved within specified limits of precision and accuracy during routine laboratory operating conditions. The EQL is generally 5 to 10 times the MDL. However, it may be nominally chosen within these guidelines to simplify data reporting. For many analytes the EQL analyte concentration is selected for the lowest non-zero standard in the calibration curve. Sample EQLs are highly matrix-dependent. The EQLs listed herein are provided for guidance and may not always be achievable. See the following information for further guidance on matrix-dependent EQLs.

^b EQLs listed for soil/sediment are based on wet weight. Normally data is reported on a dry weight basis; therefore, EQLs will be higher, based on the percent dry weight in each sample.

Other Matrices	Factor ^c
Water miscible liquid waste	50
High-concentration soil and sludge	125

^cEQL = [EQL for low soil sediment (Table 3)] X [Factor]. For non-aqueous samples, the factor is on a wet-weight basis.

TABLE 4.
BFB MASS - INTENSITY SPECIFICATIONS (4-BROMOFLUOROBENZENE)

Mass	Intensity Required (relative abundance)
50	15 to 40% of mass 95
75	30 to 60% of mass 95
95	base peak, 100% relative abundance
96	5 to 9% of mass 95
173	less than 2% of mass 174
174	greater than 50% of mass 95
175	5 to 9% of mass 174
176	greater than 95% but less than 101% of mass 174
177	5 to 9% of mass 176

TABLE 5.
CHARACTERISTIC MASSES (M/Z) FOR PURGEABLE ORGANIC COMPOUNDS

Analyte	Primary Characteristic Ion	Secondary Characteristic Ion(s)
Acetone	58	43
Acetonitrile	41	41, 40, 39
Acrolein	56	55, 58
Acrylonitrile	53	52, 51
Allyl alcohol	57	57, 58, 39
Allyl chloride	76	76, 41, 39, 78
Benzene	78	-
Benzyl chloride	91	91, 126, 65, 128
Bromoacetone	136	43, 136, 138, 93,
95 Bromobenzene	156	77, 158
Bromochloromethane	128	49, 130
Bromodichloromethane	83	85, 127
Bromoform	173	175, 254
Bromomethane	94	96
iso-Butanol	74	43
n-Butanol	56	41
2-Butanone	72	43, 72
n-Butylbenzene	91	92, 134
sec-Butylbenzene	105	134
tert-Butylbenzene	119	91, 134
Carbon disulfide	76	78
Carbon tetrachloride	117	119
Chloral hydrate	82	44, 84, 86, 111
Chloroacetonitrile	48	75
Chlorobenzene	112	77, 114
1-Chlorobutane	56	49
Chlorodibromomethane	129	208, 206
Chloroethane	64	66
2-Chloroethanol	49	49, 44, 43, 51, 80
bis-(2-chloroethyl) sulfide		109
111, 158, 160 2-Chloroethyl vinyl ether		63
65, 106 Chloroform	83	85
Chloromethane	50	52
Chloroprene	53	53, 88, 90, 51
3-Chloropropionitrile	54	54, 49, 89, 91

	2-Chlorotoluene	91	126
	4-Chlorotoluene	91	126
	1,2-Dibromo-3-chloropropane		75
155, 157			
	Dibromochloromethane	129	127
	1,2-Dibromoethane	107	109, 188
	Dibromomethane	93	95, 174
	1,2-Dichlorobenzene	146	111, 148
	1,2-Dichlorobenzene-d ₄	152	115, 150
	1,3-Dichlorobenzene	146	111, 148

TABLE 5.
(continued)

Analyte	Primary Characteristic Ion	Secondary Characteristic Ion(s)
1,4-Dichlorobenzene	146	111, 148
cis-1,4-Dichloro-2-butene		75
75, 53, 77, 124, 89		
trans-1,4-Dichloro-2-butene		53
88, 75		
Dichlorodifluoromethane	85	87
1,1-Dichloroethane	63	65, 83
1,2-Dichloroethane	62	98
1,1-Dichloroethene	96	61, 63
cis-1,2-Dichloroethene	96	61, 98
trans-1,2-Dichloroethene	96	61, 98
1,2-Dichloropropane	63	112
1,3-Dichloropropane	76	78
2,2-Dichloropropane	77	97
1,3-Dichloro-2-propanol	79	79, 43, 81, 49
1,1-Dichloropropene	75	110, 77
cis-1,3-Dichloropropene	75	77, 39
trans-1,3-Dichloropropene		75
77, 39		
1,2,3,4-Diepoxybutane	55	55, 57, 56
Diethyl ether	74	45, 59
1,4-Dioxane	88	88, 58, 43, 57
Epichlorohydrin	57	57, 49, 62, 51
Ethanol	31	45, 27, 46
Ethyl acetate	88	43, 45, 61
Ethylbenzene	91	106
Ethylene oxide	44	44, 43, 42
Ethyl methacrylate	69	69, 41, 99, 86, 114

Hexachlorobutadiene	225	223, 227
Hexachloroethane	201	166, 199, 203
2-Hexanone	43	58, 57, 100
2-Hydroxypropionitrile	44	44, 43, 42, 53
Iodomethane	142	127, 141
Isobutyl alcohol	43	43, 41, 42, 74
Isopropylbenzene	105	120
p-Isopropyltoluene	119	134, 91
Malonitrile	66	66, 39, 65, 38
Methacrylonitrile	41	41, 67, 39, 52, 66
Methyl acrylate	55	85
Methyl-t-butyl ether	73	57
Methylene chloride	84	86, 49
Methyl ethyl ketone	72	43
Methyl iodide	142	142, 127, 141
Methyl methacrylate	69	69, 41, 100, 39
4-Methyl-2-pentanone	100	43, 58, 85
Naphthalene	128	-
Nitrobenzene	123	51, 77

TABLE 5.
(continued)

Analyte	Primary Characteristic Ion	Secondary Characteristic Ion(s)
2-Nitropropane	46	-
2-Picoline	93	93, 66, 92, 78
Pentachloroethane	167	167, 130, 132, 165,
169 Propargyl alcohol	55	55, 39, 38, 53
-Propiolactone	42	42, 43, 44
Propionitrile (ethyl cyanide)		54
54, 52, 55, 40 n-Propylamine	59	59, 41, 39
n-Propylbenzene	91	120
Pyridine	79	52
Styrene	104	78
1,2,3-Trichlorobenzene	180	182, 145
1,2,4-Trichlorobenzene	180	182, 145
1,1,1,2-Tetrachloroethane		131
133, 119 1,1,2,2-Tetrachloroethane		83
131, 85 Tetrachloroethene	164	129, 131, 166

Toluene	92	91
1,1,1-Trichloroethane	97	99, 61
1,1,2-Trichloroethane	83	97, 85
Trichloroethene	95	97, 130, 132
Trichlorofluoromethane	151	101, 153
1,2,3-Trichloropropane	75	77
1,2,4-Trimethylbenzene	105	120
1,3,5-Trimethylbenzene	105	120
Vinyl acetate	43	86
Vinyl chloride	62	64
o-Xylene	106	91
m-Xylene	106	91
p-Xylene	106	91

INTERNAL STANDARDS/SURROGATES

1,4-Difluorobenzene	114	
Chlorobenzene-d ₅	117	
1,4-Dichlorobenzene-d ₄	152	115, 150
4-Bromofluorobenzene	95	174, 176
Dibromofluoromethane	113	
Dichloroethane-d ₄	102	
Toluene-d ₈	98	
Pentafluorobenzene	168	
Fluorobenzene	96	77

TABLE 6.
SINGLE LABORATORY ACCURACY AND PRECISION DATA FOR VOLATILE
ORGANIC COMPOUNDS IN WATER DETERMINED WITH A WIDE
BORE CAPILLARY COLUMN

Percent Std Analyte D	Conc. Range, _g/L	Number of Samples v	Standard	
			Recovery ^a %	Deviation ^b %
Benzene	0.1 - 10	31	97	
6.5	5.7			
Bromobenzene	0.1 - 10	30	100	

5.5	5.5			
Bromochloromethane	0.5 - 10	24		90
5.7	6.4			
Bromodichloromethane	0.1 - 10	30		95
5.7	6.1			
Bromoform	0.5 - 10	18		101
6.4	6.3			
Bromomethane	0.5 - 10	18		95
7.8	8.2			
n-Butylbenzene	0.5 - 10	18		100
7.6	7.6			
sec-Butylbenzene	0.5 - 10	16		100
7.6	7.6			
tert-Butylbenzene	0.5 - 10	18		102
7.4	7.3			
Carbon tetrachloride	0.5 - 10	24		84
7.4	8.8			
Chlorobenzene	0.1 - 10	31		98
5.8	5.9			
Chloroethane	0.5 - 10	24		89
8.0	9.0			
Chloroform	0.5 - 10	24		90
5.5	6.1			
Chloromethane	0.5 - 10	23		93
8.3	8.9			
2-Chlorotoluene	0.1 - 10	31		90
5.6	6.2			
4-Chlorotoluene	0.1 - 10	31		99
8.2	8.3			
1,2-Dibromo-3-Chloropropane	0.5 - 10	24		83
16.6	19.9			
Dibromochloromethane	0.1 - 10	31		92
6.5	7.0			
1,2-Dibromoethane	0.5 - 10	24		102
4.0	3.9			
Dibromomethane	0.5 - 10	24		100
5.6	5.6			
1,2-Dichlorobenzene	0.1 - 10	31		93
5.8	6.2			
1,3-Dichlorobenzene	0.5 - 10	24		99
6.8	6.9			
1,4-Dichlorobenzene	0.2 - 20	31		103
6.6	6.4			
Dichlorodifluoromethane	0.5 - 10	18		90
6.9	7.7			
1,1-Dichlorobenzene	0.5 - 10	24		96
5.1	5.3			
1,2-Dichlorobenzene	0.1 - 10	31		95
5.1	5.4			
1,1-Dichloroethene	0.1 - 10	34		94

6.3	6.7				
cis-1,2-Dichloroethene	0.5 - 10	18		101	
6.7	6.7				
trans-1,2-Dichloroethene	0.1 - 10	30		93	
5.2	5.6				
1,2-Dichloropropane	0.1 - 10	30		97	
5.9	6.1				
1,3-Dichloropropane	0.1 - 10	31		96	
5.7	6.0				
2,2-Dichloropropane	0.5 - 10	12		86	
14.6	16.9				
1,1-Dichloropropene	0.5 - 10	18		98	
8.7	8.9				
Ethylbenzene	0.1 - 10	31		99	
8.4	8.6				
Hexachlorobutadiene	0.5 - 10	18		100	
6.8	6.8				
Isopropylbenzene	0.5 - 10	16		101	
7.7	7.6				
p-Isopropyltoluene	0.1 - 10	23		99	
6.7	6.7				
Methylene chloride	0.1 - 10	30		95	5.0
5.3					
Naphthalene	0.1 - 100	31		104	8.6
8.2					
n-Propylbenzene	0.1 - 10	31		100	5.8
5.8					
Styrene	0.1 - 100	39		102	7.3
7.2					
1,1,1,2-Tetrachloroethane	0.5 - 10	24		90	
6.1	6.8				

TABLE 6.
(Continued)

Std Analyte	Percent	Conc. Range, _g/L	Number of Samples	Recovery ^a %	Standard Deviation ^b
1,1,2,2-Tetrachloroethane		0.1 - 10		30	91
5.7		6.3			
Tetrachloroethane		0.5 - 10	24	89	6.0
6.8					

Toluene	0.5 - 10	18	102	8.1
8.0				
1,2,3-Trichlorobenzene	0.5 - 10	18	109	9.4
8.6				
1,2,4-Trichlorobenzene	0.5 - 10	18	108	9.0
8.3				
1,1,1-Trichloroethane	0.5 - 10	18	98	7.9
8.1				
1,1,2-Trichloroethane	0.5 - 10	18	104	7.6
7.3				
Trichloroethene	0.5 - 10	24	90	6.5
7.3				
Trichlorofluoromethane	0.5 - 10	24	89	7.2
8.1				
1,2,3-Trichloropropane	0.5 - 10	16	108	15.6
14.4				
1,2,4-Trimethylbenzene	0.5 - 10	18	99	8.0
8.1				
1,3,5-Trimethylbenzene	0.5 - 10	23	92	6.8
7.4				
Vinyl chloride	0.5 - 10	18	98	6.5
6.7				
o-Xylene	0.1 - 31	18	103	7.4
7.2				
m-Xylene	0.1 - 10	31	97	6.3
6.5				
p-Xylene	0.5 - 10	18	104	8.0
7.7				

^a Recoveries were calculated using internal standard method. Internal standard was fluorobenzene.

^b Standard deviation was calculated by pooling data from three concentrations.

TABLE 7.
SINGLE LABORATORY ACCURACY AND PRECISION DATA FOR
VOLATILE ORGANIC COMPOUNDS IN WATER DETERMINED
WITH A NARROW BORE CAPILLARY COLUMN

Analyte D	Conc. _g/L	Number of Samples v	Recovery ^a %	Standard Deviation of Recovery	Percent Rel Std
Benzene 6.3	0.1	7	99	6.2	
Bromobenzene 7.6	0.5	7	97	7.4	
Bromochloromethane 6.0	0.5	7	97	5.8	
Bromodichloromethane 4.6	0.1	7	100	4.6	
Bromoform 5.3	0.5	7	101	5.4	
Bromomethane 7.2	0.5	7	99	7.1	
n-Butylbenzene 6.4	0.5	7	94	6.0	
sec-Butylbenzene 6.5	0.5	7	110	7.1	
tert-Butylbenzene 2.3	0.5	7	110	2.5	
Carbon tetrachloride 6.3	0.1	7	108	6.8	
Chlorobenzene 6.4	0.1	7	91	5.8	
Chloroethane 5.8	0.1	7	100	5.8	
Chloroform 3.0	0.1	7	105	3.2	
Chloromethane 4.7	0.5	7	101	4.7	
2-Chlorotoluene 4.6	0.5	7	99	4.6	
4-Chlorotoluene 7.3	0.5	7	96	7.0	
1,2-Dibromo-3-chloropropane 10.9	0.5	7	92	10.0	
Dibromochloromethane	0.1	7	99	5.6	

8260A - Revision 1
November 1992

5.7						
1,2-Dibromoethane	0.5	7	97		5.6	
5.8						
Dibromomethane	0.5	7	93		5.6	
6.0						
1,2-Dichlorobenzene	0.1	7	97		3.5	
3.6						
1,3-Dichlorobenzene	0.1	7	101		6.0	
5.9						
1,4-Dichlorobenzene	0.1	7	106		6.5	
6.1						
Dichlorodifluoromethane		0.1	7		99	8.8
8.9						
1,1-Dichloroethane	0.5	7	98		6.2	
6.3						
1,2-Dichloroethane	0.1	7	100		6.3	
6.3						
1,1-Dichloroethene	0.1	7	95		9.0	
9.5						
cis-1,2-Dichloroethene	0.1	7	100		3.7	
3.7						
trans-1,2-Dichloroethene		0.1	7		98	7.2
7.3						
1,2-Dichloropropane	0.5	7	96		6.0	
6.3						
1,3-Dichloropropane	0.5	7	99		5.8	
5.9						
2,2-Dichloropropane	0.5	7	99		4.9	
4.9						
1,1-Dichloropropene	0.5	7	102		7.4	
7.3						
Ethylbenzene	0.5	7	99		5.2	
5.3						
Hexachlorobutadiene	0.5	7	100		6.7	
6.7						
Isopropylbenzene	0.5	7	102		6.4	
6.3						
p-Isopropyltoluene	0.5	7	113		13.0	
11.5						
Methylene chloride	0.5	7	97		13.0	
13.4						
Naphthalene	0.5	7	98		7.2	
7.3						
n-Propylbenzene	0.5	7	99	6.6	6.7	

TABLE 7.
(Continued)

Std Analyte Dev	Conc. _g/L	Number of Samples	Recovery ^a %	Standard Deviation of Recovery	Percent Rel
Styrene 19.8	0.5	7	96	19.0	
1,1,1,2-Tetrachloroethane 4.7	0.5	7		100	4.7
1,1,2,2-Tetrachloroethane 12.0	0.5	7		100	12.0
Tetrachloroethene 5.2	0.1	7	96	5.0	
Toluene 5.9	0.5	7	100	5.9	
1,2,3-Trichlorobenzene 8.7	0.5	7	102	8.9	
1,2,4-Trichlorobenzene 17.6	0.5	7	91	16.0	
1,1,1-Trichloroethane 4.0	0.5	7	100	4.0	
1,1,2-Trichloroethane 4.8	0.5	7	102	4.9	
Trichloroethene 1.9	0.1	7	104	2.0	
Trichlorofluoromethane 4.7	0.1	7	97	4.6	
1,2,3-Trichloropropane 6.8	0.5	7	96	6.5	
1,2,4-Trimethylbenzene 6.8	0.5	7	96	6.5	
1,3,5-Trimethylbenzene 4.2	0.5	7	101	4.2	
Vinyl chloride 0.2	0.1	7	104	0.2	
o-Xylene 7.1	0.5	7	106	7.5	
m-Xylene 4.3	0.5	7	106	4.6	
p-Xylene 6.3	0.5	7	97	6.1	

^a Recoveries were calculated using internal standard method.
Internal standard was fluorobenzene.

TABLE 8.
 SURROGATE SPIKE RECOVERY LIMITS FOR WATER AND SOIL/SEDIMENT
 SAMPLES

Surrogate Compound	Low/High Water	Low/High Soil/Sediment
4-Bromofluorobenzene ^a	86-115	74-121
Dibromofluoromethane ^a	86-118	80-120
Toluene-d ₈ ^a	88-110	81-117
Dichloroethane-d ₄ ^a	80-120	80-120

^a Single laboratory data for guidance only.

TABLE 9.
 QUANTITY OF EXTRACT REQUIRED FOR ANALYSIS OF
 HIGH-CONCENTRATION SAMPLES

Approximate Concentration Range	Volume of Extract ^a
500 - 10,000 _g/kg	100 _L
1,000 - 20,000 _g/kg	50 _L
5,000 - 100,000 _g/kg	10 _L
25,000 - 500,000 _g/kg	100 _L of 1/50 dilution ^b

Calculate appropriate dilution factor for concentrations exceeding this table.

^a The volume of solvent added to 5 mL of water being purged should be kept constant. Therefore, add to the 5 mL syringe whatever volume of solvent is necessary to maintain a volume of 100 L added to the syringe.

^bDilute an aliquot of the solvent extract and then take 100 L for analysis.

TABLE 10
DIRECT INJECTION ANALYSIS OF NEW OIL AT 5 PPM

<u>Compound</u>	<u>Recovery (%)</u>	<u>%RSD</u>	<u>Blank ppm</u>	<u>Spike ppm</u>
Acetone	91	14.8	1.9	5.0
Benzene	86	21.3	0.1	0.5
n-Butanol*,**	107	27.8	0.5	5.0
iso-Butanol*,**	95	19.5	0.9	5.0
Carbon tetrachloride	86	44.7		0.0 ⑤
Carbon disulfide**	53	22.3	0.0	5.0
Chlorobenzene	81	29.3	0.0	5.0
Chloroform	84	29.3	0.0	6.0
1,4-Dichlorobenzene	98	24.9		0.0 ⑤
1,2-Dichloroethane	101	23.1	0.0	0.5
1,1-Dichloroethene	97	45.3	0.0	0.7
Diethyl ether	76	24.3	0.0	5.0
Ethyl acetate	113	27.4	0.0	5.0
Ethylbenzene	83	30.1	0.2	5.0
Hexachloroethane	71	30.3	0.0	3.0
Methylene chloride	98	45.3	0.0	5.0
Methyl ethyl ketone	79	24.6		0.4 ⑤
MIBK	93	31.4	0.0	5.0
Nitrobenzene	89	30.3	0.0	2.0
Pyridine	31	35.9	0.0	5.0
Tetrachloroethene	82	27.1	0.0	0.7
Trichlorofluoromethane	76	27.6		0.0 ⑤
1,1,2-Cl ₃ F ₃ ethane	69	29.2	0.0	5.0
Toluene	73	21.9	0.6	5.0
Trichloroethene	66	28.0	0.0	0.5
Vinyl chloride	63	35.2	0.0	0.2
o-Xylene	83	29.5	0.4	5.0
m/p-Xylene	84	29.5	0.6	10.0

* Alternate mass employed

** IS quantitation

Data taken from Reference 9.

FIGURE 1.
PURGING DEVICE

FIGURE 2.
TRAP PACKING AND CONSTRUCTION TO INCLUDE DESORB CAPABILITY

FIGURE 3.
SCHEMATIC OF PURGE-AND-TRAP DEVICE - PURGE MODE

FIGURE 4.
SCHEMATIC OF PURGE-AND-TRAP DEVICE - DESORB MODE

FIGURE 5.
GAS CHROMATOGRAM OF VOLATILE ORGANICS

FIGURE 6.
GAS CHROMATOGRAM OF VOLATILE ORGANICS

8260A - Revision 1
November 1992

FIGURE 7.
GAS CHROMATOGRAM OF VOLATILE ORGANICS

FIGURE 8.
GAS CHROMATOGRAM OF TEST MIXTURE

FIGURE 9.
LOW SOILS IMPINGER

METHOD 8260A
GAS CHROMATOGRAPHY/MASS SPECTROMETRY FOR VOLATILE ORGANICS
CAPILLARY COLUMN TECHNIQUE

8260A - Revision 1
November 1992

METHOD 8260A
(Continued)

8260A - Revision 1
November 1992